

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
 Data File : VX027259.D
 Acq On : 01 Mar 2022 19:24
 Operator : JC/MD
 Sample : N1725-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-03

Quant Time: Mar 02 00:37:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	86302	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	143178	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130847	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64100	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51259	49.454	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.900%
35) Dibromofluoromethane	5.391	113	44269	49.812	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.620%
50) Toluene-d8	8.653	98	166826	50.202	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.400%
62) 4-Bromofluorobenzene	11.085	95	62745	49.190	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.380%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1166	2.562	ug/l #	85
20) Methylene Chloride	2.794	84	270	0.286	ug/l #	78
40) Benzene	6.050	78	4861	1.282	ug/l	95
68) m/p-Xylenes	10.305	106	880	0.489	ug/l	95
69) o-Xylene	10.646	106	648	0.359	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	2856	0.744	ug/l	92
95) Naphthalene	13.780	128	1603343	355.968	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
Data File : VX027259.D
Acq On : 01 Mar 2022 19:24
Operator : JC/MD
Sample : N1725-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-03

Quant Time: Mar 02 00:37:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

