

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
 Data File : VX027261.D
 Acq On : 01 Mar 2022 20:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 02 00:38:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	50.000	43.702	12.6	83	0.00
3 P	Chloromethane	50.000	41.279	17.4	76	0.00
4 C	Vinyl Chloride	50.000	40.886	18.2#	74	0.00
5 T	Bromomethane	50.000	26.132	47.7#	54	0.00
6 T	Chloroethane	50.000	41.635	16.7	78	0.00
7 T	Trichlorofluoromethane	50.000	46.369	7.3	86	0.00
8 T	Diethyl Ether	50.000	44.563	10.9	80	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.480	5.0	87	0.00
10 T	Methyl Iodide	50.000	26.179	47.6#	46	0.00
11 T	Tert butyl alcohol	250.000	226.327	9.5	85	0.00
12 CM	1,1-Dichloroethene	50.000	46.750	6.5#	86	0.00
13 T	Acrolein	250.000	210.163	15.9	81	0.00
14 T	Allyl chloride	50.000	45.854	8.3	83	0.00
15 T	Acrylonitrile	250.000	245.743	1.7	88	0.00
16 T	Acetone	250.000	229.595	8.2	88	0.00
17 T	Carbon Disulfide	50.000	44.643	10.7	82	0.00
18 T	Methyl Acetate	50.000	46.776	6.4	85	0.00
19 T	Methyl tert-butyl Ether	50.000	48.119	3.8	87	0.00
20 T	Methylene Chloride	50.000	47.540	4.9	88	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.046	5.9	86	0.00
22 T	Diisopropyl ether	50.000	47.365	5.3	85	0.00
23 T	Vinyl Acetate	250.000	245.367	1.9	86	0.00
24 P	1,1-Dichloroethane	50.000	47.687	4.6	87	0.00
25 T	2-Butanone	250.000	240.666	3.7	87	0.00
26 T	2,2-Dichloropropane	50.000	46.136	7.7	83	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.351	3.3	87	0.00
28 T	Bromochloromethane	50.000	45.030	9.9	80	0.00
29 T	Tetrahydrofuran	250.000	234.788	6.1	86	0.00
30 C	Chloroform	50.000	47.783	4.4#	88	0.00
31 T	Cyclohexane	50.000	45.331	9.3	83	0.00
32 T	1,1,1-Trichloroethane	50.000	47.994	4.0	86	0.00
33 S	1,2-Dichloroethane-d4	50.000	42.450	15.1	81	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	95	0.00
35 S	Dibromofluoromethane	50.000	42.619	14.8	83	0.00
36 T	1,1-Dichloropropene	50.000	46.052	7.9	85	0.00
37 T	Ethyl Acetate	50.000	47.400	5.2	89	0.00
38 T	Carbon Tetrachloride	50.000	47.274	5.5	86	0.00
39 T	Methylcyclohexane	50.000	46.101	7.8	85	0.00
40 TM	Benzene	50.000	46.997	6.0	87	0.00
41 T	Methacrylonitrile	50.000	49.609	0.8	91	0.00
42 TM	1,2-Dichloroethane	50.000	48.265	3.5	89	0.00
43 T	Isopropyl Acetate	50.000	48.107	3.8	87	0.00
44 TM	Trichloroethene	50.000	46.218	7.6	85	0.00
45 C	1,2-Dichloropropane	50.000	48.846	2.3#	88	0.00
46 T	Dibromomethane	50.000	49.255	1.5	89	0.00
47 T	Bromodichloromethane	50.000	49.216	1.6	91	0.00
48 T	Methyl methacrylate	50.000	49.951	0.1	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\X030122\
 Data File : VX027261.D
 Acq On : 01 Mar 2022 20:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 02 00:38:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	966.684	3.3	88	0.00
50 S	Toluene-d8	50.000	41.764	16.5	81	0.00
51 T	4-Methyl-2-Pentanone	250.000	244.262	2.3	89	0.00
52 CM	Toluene	50.000	49.044	1.9#	90	0.00
53 T	t-1,3-Dichloropropene	50.000	44.359	11.3	88	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.278	-0.6	89	0.00
55 T	1,1,2-Trichloroethane	50.000	49.558	0.9	90	0.00
56 T	Ethyl methacrylate	50.000	49.153	1.7	87	0.00
57 T	1,3-Dichloropropane	50.000	49.490	1.0	89	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	248.884	0.4	89	0.00
59 T	2-Hexanone	250.000	244.460	2.2	88	0.00
60 T	Dibromochloromethane	50.000	51.235	-2.5	91	0.00
61 T	1,2-Dibromoethane	50.000	48.586	2.8	87	0.00
62 S	4-Bromofluorobenzene	50.000	41.748	16.5	80	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	47.438	5.1	88	0.00
65 PM	Chlorobenzene	50.000	47.623	4.8	88	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.812	2.4	89	0.00
67 C	Ethyl Benzene	50.000	47.818	4.4#	88	0.00
68 T	m/p-Xylenes	100.000	96.897	3.1	87	0.00
69 T	o-Xylene	50.000	48.282	3.4	89	0.00
70 T	Styrene	50.000	49.196	1.6	88	0.00
71 P	Bromoform	50.000	44.209	11.6	88	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	94	0.00
73 T	Isopropylbenzene	50.000	48.224	3.6	88	0.00
74 T	N-ethyl acetate	50.000	49.822	0.4	87	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.939	-3.9	98	0.00
76 T	1,2,3-Trichloropropane	50.000	48.717	2.6	89	0.00
77 T	Bromobenzene	50.000	47.934	4.1	88	0.00
78 T	n-propylbenzene	50.000	50.558	-1.1	92	0.00
79 T	2-Chlorotoluene	50.000	47.528	4.9	88	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.021	4.0	87	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	38.110	23.8	74	0.00
82 T	4-Chlorotoluene	50.000	47.419	5.2	87	0.00
83 T	tert-Butylbenzene	50.000	49.106	1.8	90	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.029	-0.1	90	0.00
85 T	sec-Butylbenzene	50.000	48.092	3.8	87	0.00
86 T	p-Isopropyltoluene	50.000	49.989	0.0	90	0.00
87 T	1,3-Dichlorobenzene	50.000	47.461	5.1	87	0.00
88 T	1,4-Dichlorobenzene	50.000	46.485	7.0	87	0.00
89 T	n-Butylbenzene	50.000	47.904	4.2	85	0.00
90 T	Hexachloroethane	50.000	43.440	13.1	77	0.00
91 T	1,2-Dichlorobenzene	50.000	47.681	4.6	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.350	1.3	88	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.789	4.4	86	0.00
94 T	Hexachlorobutadiene	50.000	46.355	7.3	84	0.00
95 T	Naphthalene	50.000	51.100	-2.2	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030122\
Data File : VX027261.D
Acq On : 01 Mar 2022 20:11
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 02 00:38:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.189	3.6	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6