

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030123\
 Data File : VX034281.D
 Acq On : 01 Mar 2023 10:22
 Operator : JC/MD
 Sample : VX0301MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0301MBL01

Quant Time: Mar 02 00:38:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	285046	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	476309	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	419322	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	189245	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	172919	45.044	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.080%
35) Dibromofluoromethane	5.379	113	149057	48.064	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.120%
50) Toluene-d8	8.647	98	569986	51.191	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.380%
62) 4-Bromofluorobenzene	11.079	95	193235	45.371	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030123\
Data File : VX034281.D
Acq On : 01 Mar 2023 10:22
Operator : JC/MD
Sample : VX0301MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0301MBL01

Quant Time: Mar 02 00:38:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

