

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030123\
 Data File : VX034286.D
 Acq On : 01 Mar 2023 12:33
 Operator : JC/MD
 Sample : PB151127TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151127TB

Quant Time: Mar 02 00:41:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	275407	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	453133	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	405883	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	174248	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	163615	44.112	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.220%
35) Dibromofluoromethane	5.385	113	139742	47.365	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.720%
50) Toluene-d8	8.647	98	545486	51.496	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.000%
62) 4-Bromofluorobenzene	11.079	95	184598	45.560	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.120%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	93919	49.087	ug/l	94
20) Methylene Chloride	2.794	84	24484	7.102	ug/l	96
37) Ethyl Acetate	4.715	43	63030	12.926	ug/l	100
43) Isopropyl Acetate	6.336	43	258138	29.998	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030123\
Data File : VX034286.D
Acq On : 01 Mar 2023 12:33
Operator : JC/MD
Sample : PB151127TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151127TB

Quant Time: Mar 02 00:41:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

