

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030123\
 Data File : VX034287.D
 Acq On : 01 Mar 2023 12:56
 Operator : JC/MD
 Sample : 01739-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TCLP-VOC-1

Quant Time: Mar 02 00:41:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270687	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	446697	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	392205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	166847	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	159118	43.648	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.300%
35) Dibromofluoromethane	5.385	113	139753	48.051	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.100%
50) Toluene-d8	8.647	98	533050	51.047	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.100%
62) 4-Bromofluorobenzene	11.079	95	178556	44.704	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.400%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	51875	27.585	ug/l	96
20) Methylene Chloride	2.794	84	9455	2.790	ug/l	97
37) Ethyl Acetate	4.721	43	75722	15.752	ug/l	99
43) Isopropyl Acetate	6.342	43	303986	35.835	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030123\
Data File : VX034287.D
Acq On : 01 Mar 2023 12:56
Operator : JC/MD
Sample : 01739-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TCLP-VOC-1

Quant Time: Mar 02 00:41:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

