

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030123\
 Data File : VX034289.D
 Acq On : 01 Mar 2023 13:44
 Operator : JC/MD
 Sample : 01739-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TCLP-VOC-3

Quant Time: Mar 02 00:42:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	267846	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	440824	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	396605	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	170298	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	160706	44.551	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.100%
35) Dibromofluoromethane	5.385	113	138928	48.404	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.800%
50) Toluene-d8	8.647	98	529227	51.356	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.720%
62) 4-Bromofluorobenzene	11.079	95	181484	46.042	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.080%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	54218	29.137	ug/l #	88
18) Methyl Acetate	2.703	43	19254	3.344	ug/l	98
20) Methylene Chloride	2.794	84	9457	2.821	ug/l	88
37) Ethyl Acetate	4.721	43	73084	15.406	ug/l	99
43) Isopropyl Acetate	6.342	43	301968	36.071	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030123\
Data File : VX034289.D
Acq On : 01 Mar 2023 13:44
Operator : JC/MD
Sample : 01739-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TCLP-VOC-3

Quant Time: Mar 02 00:42:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

