

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX030123\  
 Data File : VX034280.D  
 Acq On : 01 Mar 2023 09:43  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 03/02/2023  
 Supervised By :Mahesh Dadoda 03/02/2023

Quant Time: Mar 02 00:38:02 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X021723W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 17 12:33:35 2023  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                        |                |      |                     |         |        |          |
| Internal Standards           |                |      |                     |         |        |          |
| 1) Pentafluorobenzene        | 5.550          | 168  | 348258              | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.757          | 114  | 568385              | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 525881              | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 266228              | 50.000  | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |                     |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.952          | 65   | 218491              | 46.584  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 78 - 117 |      | Recovery = 93.160%  |         |        |          |
| 35) Dibromofluoromethane     | 5.385          | 113  | 185764              | 50.197  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = 100.400% |         |        |          |
| 50) Toluene-d8               | 8.647          | 98   | 689916              | 51.924  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 92 - 112 |      | Recovery = 103.840% |         |        |          |
| 62) 4-Bromofluorobenzene     | 11.079         | 95   | 282150              | 55.516  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = 111.040% |         |        |          |
| Target Compounds             |                |      |                     |         |        |          |
|                              |                |      |                     |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.166          | 85   | 157263              | 52.259  | ug/l   | 99       |
| 3) Chloromethane             | 1.294          | 50   | 219851              | 50.767  | ug/l   | 98       |
| 4) Vinyl Chloride            | 1.374          | 62   | 222513              | 56.419  | ug/l   | 98       |
| 5) Bromomethane              | 1.599          | 94   | 116186              | 81.919  | ug/l   | 99       |
| 6) Chloroethane              | 1.685          | 64   | 122490              | 61.779  | ug/l   | 98       |
| 7) Trichlorofluoromethane    | 1.886          | 101  | 317841              | 59.569  | ug/l   | 98       |
| 8) Diethyl Ether             | 2.130          | 74   | 116527              | 54.208  | ug/l   | 97       |
| 9) 1,1,2-Trichlorotrifluo... | 2.325          | 101  | 178704              | 47.557  | ug/l   | 96       |
| 10) Methyl Iodide            | 2.453          | 142  | 225515              | 49.291  | ug/l   | 94       |
| 11) Tert butyl alcohol       | 2.953          | 59   | 202658              | 193.187 | ug/l # | 84       |
| 12) 1,1-Dichloroethene       | 2.319          | 96   | 178307              | 47.337  | ug/l   | 93       |
| 13) Acrolein                 | 2.233          | 56   | 245932              | 513.578 | ug/l   | 100      |
| 14) Allyl chloride           | 2.660          | 41   | 352879              | 44.163  | ug/l   | 90       |
| 15) Acrylonitrile            | 3.056          | 53   | 525064              | 229.438 | ug/l   | 99       |
| 16) Acetone                  | 2.373          | 43   | 602276              | 248.932 | ug/l   | 94       |
| 17) Carbon Disulfide         | 2.514          | 76   | 512896              | 47.360  | ug/l   | 98       |
| 18) Methyl Acetate           | 2.697          | 43   | 334101              | 44.628  | ug/l   | 100      |
| 19) Methyl tert-butyl Ether  | 3.111          | 73   | 625549              | 47.053  | ug/l   | 97       |
| 20) Methylene Chloride       | 2.788          | 84   | 201615              | 46.249  | ug/l   | 99       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 196655              | 49.823  | ug/l   | 98       |
| 22) Diisopropyl ether        | 3.751          | 45   | 719271              | 51.347  | ug/l   | 93       |
| 23) Vinyl Acetate            | 3.715          | 43   | 2790434             | 248.633 | ug/l   | 99       |
| 24) 1,1-Dichloroethane       | 3.605          | 63   | 368157              | 47.874  | ug/l   | 99       |
| 25) 2-Butanone               | 4.544          | 43   | 825465              | 235.798 | ug/l   | 99       |
| 26) 2,2-Dichloropropane      | 4.471          | 77   | 327618              | 47.028  | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 4.483          | 96   | 232800              | 49.807  | ug/l   | 96       |
| 28) Bromochloromethane       | 4.897          | 49   | 161745              | 47.149  | ug/l   | 88       |
| 29) Tetrahydrofuran          | 4.995          | 42   | 493567              | 231.115 | ug/l   | 98       |
| 30) Chloroform               | 5.092          | 83   | 368533              | 49.907  | ug/l   | 99       |
| 31) Cyclohexane              | 5.464          | 56   | 352120              | 50.285  | ug/l   | 95       |
| 32) 1,1,1-Trichloroethane    | 5.379          | 97   | 326995              | 49.398  | ug/l   | 99       |
| 36) 1,1-Dichloropropene      | 5.690          | 75   | 282140              | 51.551  | ug/l   | 99       |
| 37) Ethyl Acetate            | 4.708          | 43   | 303301              | 49.586  | ug/l   | 99       |
| 38) Carbon Tetrachloride     | 5.678          | 117  | 288071              | 51.150  | ug/l   | 97       |
| 39) Methylcyclohexane        | 7.379          | 83   | 361872              | 51.379  | ug/l   | 98       |
| 40) Benzene                  | 6.031          | 78   | 833874              | 51.413  | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX030123\  
 Data File : VX034280.D  
 Acq On : 01 Mar 2023 09:43  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 03/02/2023  
 Supervised By :Mahesh Dadoda 03/02/2023

Quant Time: Mar 02 00:38:02 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X021723W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 17 12:33:35 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.916  | 41   | 170083   | 49.908  | ug/l   | 99       |
| 42) 1,2-Dichloroethane        | 6.080  | 62   | 292032   | 51.194  | ug/l   | 99       |
| 43) Isopropyl Acetate         | 6.330  | 43   | 525463   | 48.682  | ug/l   | 99       |
| 44) Trichloroethene           | 7.123  | 130  | 217516   | 52.260  | ug/l   | 96       |
| 45) 1,2-Dichloropropane       | 7.427  | 63   | 218345   | 50.570  | ug/l   | 98       |
| 46) Dibromomethane            | 7.580  | 93   | 143816   | 50.612  | ug/l   | 89       |
| 47) Bromodichloromethane      | 7.818  | 83   | 296486   | 50.842  | ug/l   | 97       |
| 48) Methyl methacrylate       | 7.690  | 41   | 249428   | 48.687  | ug/l   | 98       |
| 49) 1,4-Dioxane               | 7.653  | 88   | 92175    | 902.032 | ug/l # | 99       |
| 51) 4-Methyl-2-Pentanone      | 8.567  | 43   | 1526046  | 256.974 | ug/l   | 99       |
| 52) Toluene                   | 8.714  | 92   | 534301   | 54.367  | ug/l   | 98       |
| 53) t-1,3-Dichloropropene     | 8.976  | 75   | 352992   | 52.588  | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 8.366  | 75   | 365992   | 50.931  | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 210981   | 53.186  | ug/l   | 99       |
| 56) Ethyl methacrylate        | 9.116  | 69   | 351553   | 52.388  | ug/l   | 96       |
| 57) 1,3-Dichloropropane       | 9.305  | 76   | 361842   | 52.428  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 8.238  | 63   | 706530   | 243.048 | ug/l   | 95       |
| 59) 2-Hexanone                | 9.427  | 43   | 1202515  | 265.703 | ug/l   | 98       |
| 60) Dibromochloromethane      | 9.518  | 129  | 236209   | 53.891  | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 226039   | 53.879  | ug/l   | 99       |
| 64) Tetrachloroethene         | 9.269  | 164  | 186463   | 50.239  | ug/l   | 93       |
| 65) Chlorobenzene             | 10.079 | 112  | 568666   | 51.876  | ug/l   | 98       |
| 66) 1,1,1,2-Tetrachloroethane | 10.159 | 131  | 214605   | 50.571  | ug/l   | 98       |
| 67) Ethyl Benzene             | 10.195 | 91   | 1038654  | 52.740  | ug/l   | 98       |
| 68) m/p-Xylenes               | 10.299 | 106  | 794229   | 105.384 | ug/l   | 94       |
| 69) o-Xylene                  | 10.640 | 106  | 396347   | 52.825  | ug/l   | 94       |
| 70) Styrene                   | 10.652 | 104  | 654886   | 53.667  | ug/l   | 96       |
| 71) Bromoform                 | 10.799 | 173  | 177137   | 51.138  | ug/l   | 100      |
| 73) Isopropylbenzene          | 10.963 | 105  | 1026159  | 50.575  | ug/l   | 100      |
| 74) N-amyl acetate            | 10.841 | 43   | 493129   | 48.377  | ug/l   | 96       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 329351   | 47.459  | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 11.238 | 75   | 278897m  | 41.632  | ug/l   |          |
| 77) Bromobenzene              | 11.195 | 156  | 244028   | 49.977  | ug/l   | 83       |
| 78) n-propylbenzene           | 11.305 | 91   | 1220864  | 51.737  | ug/l   | 98       |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 737137   | 50.944  | ug/l   | 94       |
| 80) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 864649   | 50.686  | ug/l   | 96       |
| 81) trans-1,4-Dichloro-2-b... | 11.018 | 75   | 119759   | 45.925  | ug/l   | 94       |
| 82) 4-Chlorotoluene           | 11.451 | 91   | 841379   | 51.053  | ug/l   | 95       |
| 83) tert-Butylbenzene         | 11.713 | 119  | 862255   | 49.353  | ug/l   | 93       |
| 84) 1,2,4-Trimethylbenzene    | 11.750 | 105  | 875272   | 50.746  | ug/l   | 97       |
| 85) sec-Butylbenzene          | 11.890 | 105  | 1092721  | 51.503  | ug/l   | 98       |
| 86) p-Isopropyltoluene        | 12.006 | 119  | 913919   | 52.049  | ug/l   | 96       |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 468918   | 51.207  | ug/l   | 97       |
| 88) 1,4-Dichlorobenzene       | 12.042 | 146  | 468257   | 50.697  | ug/l   | 96       |
| 89) n-Butylbenzene            | 12.329 | 91   | 829952   | 53.069  | ug/l   | 98       |
| 90) Hexachloroethane          | 12.536 | 117  | 182235   | 49.638  | ug/l   | 79       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 452222   | 50.474  | ug/l   | 97       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.939 | 75   | 73071    | 44.155  | ug/l   | 77       |
| 93) 1,2,4-Trichlorobenzene    | 13.585 | 180  | 291367   | 50.654  | ug/l   | 98       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 123015   | 51.588  | ug/l   | 99       |
| 95) Naphthalene               | 13.774 | 128  | 929798   | 47.984  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 283392   | 49.625  | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX030123\  
Data File : VX034280.D  
Acq On : 01 Mar 2023 09:43  
Operator : JC/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023

Quant Time: Mar 02 00:38:02 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X021723W.M  
Quant Title : SW846 8260  
QLast Update : Fri Feb 17 12:33:35 2023  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|


-----  
(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.550 min Scan# 71  
Delta R.T. -0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:168 Resp: 348258  
Ion Ratio Lower Upper  
168 100  
99 55.4 39.6 59.4

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#2  
Dichlorodifluoromethane  
Concen: 52.259 ug/l  
RT: 1.166 min Scan# 13  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 85 Resp: 157263  
Ion Ratio Lower Upper  
85 100  
87 32.4 15.8 47.4





#3  
Chloromethane  
Concen: 50.767 ug/l  
RT: 1.294 min Scan# 34  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 50 Resp: 21985  
Ion Ratio Lower Upper  
50 100  
52 32.5 25.2 37.8

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#4  
Vinyl Chloride  
Concen: 56.419 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 62 Resp: 222513  
Ion Ratio Lower Upper  
62 100  
64 31.6 26.2 39.4





#5  
Bromomethane  
Concen: 81.919 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 94 Resp: 116180  
Ion Ratio Lower Upper  
94 100  
96 92.4 74.3 111.5

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023






#6  
Chloroethane  
Concen: 61.779 ug/l  
RT: 1.685 min Scan# 98  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43



Tgt Ion: 64 Resp: 122490  
Ion Ratio Lower Upper  
64 100  
66 31.4 26.1 39.1



Abundance Scan 131 (1.886 min): VX034012.D\data.ms (-12



#7

Trichlorofluoromethane

Concen: 59.569 ug/l

RT: 1.886 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX034280.D

Acq: 01 Mar 2023 09:43

Tgt Ion:101 Resp: 31784

Ion Ratio Lower Upper

101 100

103 64.4 50.6 75.8

Instrument :

MSVOA\_X

ClientSampleId :


VSTDCCC050

Manual Integrations




APPROVED

Reviewed By :Krupa Patel 03/02/2023

Supervised By :Mahesh Dadoda 03/02/2023



Abundance Scan 171 (2.130 min): VX034012.D\data.ms (-16



#8

Diethyl Ether

Concen: 54.208 ug/l

RT: 2.130 min Scan# 171

Delta R.T. 0.000 min

Lab File: VX034280.D


Acq: 01 Mar 2023 09:43


Tgt Ion: 74 Resp: 116527

Ion Ratio Lower Upper

74 100

45 100.1 51.5 154.5









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 47.557 ug/l  
RT: 2.325 min Scan# 204  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion:101 Resp: 178704  
Ion Ratio Lower Upper  
101 100  
85 44.6 36.7 55.1  
151 78.0 66.5 99.7




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#10  
Methyl Iodide  
Concen: 49.291 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Tgt Ion:142 Resp: 225515  
Ion Ratio Lower Upper  
142 100  
127 41.8 37.3 55.9  
141 14.4 11.5 17.3









#11  
Tert butyl alcohol  
Concen: 193.187 ug/l  
RT: 2.953 min Scan# 307  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion: 59 Resp: 202658  
Ion Ratio Lower Upper  
59 100  
57 10.0 3.7 5.5




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#12  
1,1-Dichloroethene  
Concen: 47.337 ug/l  
RT: 2.319 min Scan# 202  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43


Tgt Ion: 96 Resp: 178307  
Ion Ratio Lower Upper  
96 100  
61 155.5 134.2 201.4  
98 63.0 50.2 75.2





#13  
Acrolein  
Concen: 513.578 ug/l  
RT: 2.233 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 56 Resp: 24593  
Ion Ratio Lower Upper  
56 100  
55 70.2 56.3 84.5

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#14  
Allyl chloride  
Concen: 44.163 ug/l  
RT: 2.660 min Scan# 258  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 41 Resp: 352879  
Ion Ratio Lower Upper  
41 100  
39 63.1 59.0 88.4  
76 33.1 24.6 37.0





#15  
Acrylonitrile  
Concen: 229.438 ug/l  
RT: 3.056 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 53 Resp: 525064  
Ion Ratio Lower Upper  
53 100  
52 81.8 65.7 98.5  
51 35.3 29.0 43.4

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#16  
Acetone  
Concen: 248.932 ug/l  
RT: 2.373 min Scan# 211  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 43 Resp: 602276  
Ion Ratio Lower Upper  
43 100  
58 31.3 22.6 33.8





#17  
Carbon Disulfide  
Concen: 47.360 ug/l  
RT: 2.514 min Scan# 21  
Delta R.T. 0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 76 Resp: 512890  
Ion Ratio Lower Upper  
76 100  
78 9.5 7.0 10.6

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#18  
Methyl Acetate  
Concen: 44.628 ug/l  
RT: 2.697 min Scan# 264  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 43 Resp: 334101  
Ion Ratio Lower Upper  
43 100  
74 22.9 18.2 27.4





#19  
Methyl tert-butyl Ether  
Concen: 47.053 ug/l  
RT: 3.111 min Scan# 311  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 73 Resp: 625549  
Ion Ratio Lower Upper  
73 100  
57 22.9 17.4 26.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#20  
Methylene Chloride  
Concen: 46.249 ug/l  
RT: 2.788 min Scan# 279  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43



Tgt Ion: 84 Resp: 201615  
Ion Ratio Lower Upper  
84 100  
49 127.1 101.1 151.7  
51 39.3 31.0 46.6  
86 64.4 52.6 79.0






#21  
trans-1,2-Dichloroethene  
Concen: 49.823 ug/l  
RT: 3.093 min Scan# 31  
Delta R.T. 0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050






Tgt Ion: 96 Resp: 196658  
Ion Ratio Lower Upper  
96 100  
61 136.8 106.7 160.1  
98 61.8 50.0 75.0

Manual Integrations


APPROVED

Reviewed By :Krupa Patel 03/02/2023



Supervised By :Mahesh Dadoda 03/02/2023




#22  
Diisopropyl ether  
Concen: 51.347 ug/l  
RT: 3.751 min Scan# 437  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 45 Resp: 719271  
Ion Ratio Lower Upper  
45 100  
43 89.0 64.6 97.0  
87 25.3 21.9 32.9  
59 10.8 9.2 13.8





#23  
 Vinyl Acetate  
 Concen: 248.633 ug/l  
 RT: 3.715 min Scan# 413  
 Delta R.T. 0.000 min  
 Lab File: VX034280.D  
 Acq: 01 Mar 2023 09:43




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050




Tgt Ion: 43 Resp: 2790434  
 Ion Ratio Lower Upper  
 43 100  
 86 10.2 8.4 12.6

Manual Integrations  
 APPROVED



Reviewed By :Krupa Patel 03/02/2023  
 Supervised By :Mahesh Dadoda 03/02/2023




#24  
 1,1-Dichloroethane  
 Concen: 47.874 ug/l  
 RT: 3.605 min Scan# 413  
 Delta R.T. 0.000 min  
 Lab File: VX034280.D  
 Acq: 01 Mar 2023 09:43




Tgt Ion: 63 Resp: 368157  
 Ion Ratio Lower Upper  
 63 100  
 98 7.0 3.8 11.2  
 100 4.5 2.4 7.2





#25  
2-Butanone  
Concen: 235.798 ug/l  
RT: 4.544 min Scan# 50  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 43 Resp: 82546  
Ion Ratio Lower Upper  
43 100  
72 24.7 19.4 29.2

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#26  
2,2-Dichloropropane  
Concen: 47.028 ug/l  
RT: 4.471 min Scan# 555  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 77 Resp: 327618  
Ion Ratio Lower Upper  
77 100  
97 23.1 11.9 35.5





#27  
cis-1,2-Dichloroethene  
Concen: 49.807 ug/l  
RT: 4.483 min Scan# 511  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 96 Resp: 232800  
Ion Ratio Lower Upper  
96 100  
61 145.7 0.0 277.6  
98 63.9 0.0 128.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#28  
Bromochloromethane  
Concen: 47.149 ug/l  
RT: 4.897 min Scan# 625  
Delta R.T. 0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43



Tgt Ion: 49 Resp: 161745  
Ion Ratio Lower Upper  
49 100  
129 1.8 0.0 4.6  
130 73.9 68.5 102.7





#29

Tetrahydrofuran

Concen: 231.115 ug/l

RT: 4.995 min Scan# 64

Delta R.T. 0.000 min

Lab File: VX034280.D

Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 49356

Ion Ratio Lower Upper

42 100

72 42.7 34.8 52.2





71 39.0 32.2 48.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 03/02/2023

Supervised By :Mahesh Dadoda 03/02/2023



#30

Chloroform

Concen: 49.907 ug/l

RT: 5.092 min Scan# 657

Delta R.T. 0.000 min

Lab File: VX034280.D




Acq: 01 Mar 2023 09:43


Tgt Ion: 83 Resp: 368533

Ion Ratio Lower Upper

83 100


85 65.9 52.1 78.1





#31  
Cyclohexane  
Concen: 50.285 ug/l  
RT: 5.464 min Scan# 719  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 56 Resp: 352120  
Ion Ratio Lower Upper  
56 100  
69 29.7 24.6 37.0  
84 83.1 71.0 106.6

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#32  
1,1,1-Trichloroethane  
Concen: 49.398 ug/l  
RT: 5.379 min Scan# 704  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 97 Resp: 326995  
Ion Ratio Lower Upper  
97 100  
99 64.7 52.3 78.5  
61 44.6 35.2 52.8





#33  
1,2-Dichloroethane-d4  
Concen: 46.584 ug/l  
RT: 5.952 min Scan# 798  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050



Tgt Ion: 65 Resp: 21849  
Ion Ratio Lower Upper  
65 100  
67 53.4 0.0 104.8




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.757 min Scan# 930  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43


Tgt Ion:114 Resp: 568385  
Ion Ratio Lower Upper  
114 100  
63 21.4 0.0 37.4  
88 16.3 0.0 30.6





#35  
Dibromofluoromethane  
Concen: 50.197 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:113 Resp: 185764  
Ion Ratio Lower Upper  
113 100  
111 102.8 83.0 124.6  
192 19.5 19.0 28.4

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#36  
1,1-Dichloropropene  
Concen: 51.551 ug/l  
RT: 5.690 min Scan# 755  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 75 Resp: 282140  
Ion Ratio Lower Upper  
75 100  
110 36.4 18.8 56.4  
77 31.2 25.1 37.7





#37  
Ethyl Acetate  
Concen: 49.586 ug/l  
RT: 4.708 min Scan# 594  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050



Tgt Ion: 43 Resp: 30330  
Ion Ratio Lower Upper  
43 100  
61 13.3 11.1 16.7  
70 10.7 8.5 12.7




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#38  
Carbon Tetrachloride  
Concen: 51.150 ug/l  
RT: 5.678 min Scan# 753  
Delta R.T. 0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43


Tgt Ion:117 Resp: 288071  
Ion Ratio Lower Upper  
117 100  
119 94.5 78.0 117.0  
121 31.1 24.4 36.6





#39  
Methylcyclohexane  
Concen: 51.379 ug/l  
RT: 7.379 min Scan# 1032  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 83 Resp: 361872  
Ion Ratio Lower Upper  
83 100  
55 78.7 61.8 92.6  
98 50.8 39.3 58.9

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#40  
Benzene  
Concen: 51.413 ug/l  
RT: 6.031 min Scan# 811  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43



Tgt Ion: 78 Resp: 833874  
Ion Ratio Lower Upper  
78 100  
77 23.6 19.0 28.4






#41  
Methacrylonitrile  
Concen: 49.908 ug/l  
RT: 4.916 min Scan# 61  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050






Tgt Ion: 41 Resp: 17008  
Ion Ratio Lower Upper  
41 100  
39 53.5 43.8 65.6  
67 68.4 54.8 82.2  
52 27.2 23.0 34.6

Manual Integrations


APPROVED

Reviewed By :Krupa Patel 03/02/2023



Supervised By :Mahesh Dadoda 03/02/2023




#42  
1,2-Dichloroethane  
Concen: 51.194 ug/l  
RT: 6.080 min Scan# 819  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 62 Resp: 292032  
Ion Ratio Lower Upper  
62 100  
98 9.6 0.0 20.2





#43  
Isopropyl Acetate  
Concen: 48.682 ug/l  
RT: 6.330 min Scan# 861  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050



Tgt Ion: 43 Resp: 52546  
Ion Ratio Lower Upper  
43 100  
61 19.8 15.7 23.5  
87 12.5 10.6 15.8




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#44  
Trichloroethene  
Concen: 52.260 ug/l  
RT: 7.123 min Scan# 990  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Tgt Ion:130 Resp: 217516  
Ion Ratio Lower Upper  
130 100  
95 95.3 0.0 182.6





#45

1,2-Dichloropropane

Concen: 50.570 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX034280.D

Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 21834

Ion Ratio Lower Upper

63 100





65 31.1 24.2 36.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 03/02/2023

Supervised By :Mahesh Dadoda 03/02/2023



#46

Dibromomethane

Concen: 50.612 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.000 min

Lab File: VX034280.D

Acq: 01 Mar 2023 09:43




Tgt Ion: 93 Resp: 143816


Ion Ratio Lower Upper

93 100

95 84.3 67.6 101.4


174 100.6 97.4 146.0





#47  
Bromodichloromethane  
Concen: 50.842 ug/l  
RT: 7.818 min Scan# 1104  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 83 Resp: 296480  
Ion Ratio Lower Upper  
83 100  
85 64.7 49.8 74.6  
127 7.9 7.0 10.4

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#48  
Methyl methacrylate  
Concen: 48.687 ug/l  
RT: 7.690 min Scan# 1083  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 41 Resp: 249428  
Ion Ratio Lower Upper  
41 100  
69 79.4 63.1 94.7  
39 57.6 43.2 64.8





#49  
1,4-Dioxane  
Concen: 902.032 ug/l  
RT: 7.653 min Scan# 1079  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050



Tgt Ion: 88 Resp: 92175  
Ion Ratio Lower Upper  
88 100  
43 33.9 0.0 0.0  
58 71.1 56.5 84.7




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#50  
Toluene-d8  
Concen: 51.924 ug/l  
RT: 8.647 min Scan# 1240  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Tgt Ion: 98 Resp: 689916  
Ion Ratio Lower Upper  
98 100  
100 66.1 53.6 80.4









#51  
4-Methyl-2-Pentanone  
Concen: 256.974 ug/l  
RT: 8.567 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion: 43 Resp: 152604  
Ion Ratio Lower Upper  
43 100  
58 38.1 31.2 46.8




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#52  
Toluene  
Concen: 54.367 ug/l  
RT: 8.714 min Scan# 1251  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Tgt Ion: 92 Resp: 534301  
Ion Ratio Lower Upper  
92 100  
91 172.1 135.5 203.3





#53

t-1,3-Dichloropropene

Concen: 52.588 ug/l

RT: 8.976 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX034280.D

Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 75 Resp: 352992

Ion Ratio Lower Upper

75 100





77 31.6 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 03/02/2023

Supervised By :Mahesh Dadoda 03/02/2023



#54

cis-1,3-Dichloropropene

Concen: 50.931 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX034280.D

Acq: 01 Mar 2023 09:43




Tgt Ion: 75 Resp: 365992


Ion Ratio Lower Upper

75 100

77 31.4 25.0 37.6


39 50.6 38.6 58.0





#55  
1,1,2-Trichloroethane  
Concen: 53.186 ug/l  
RT: 9.153 min Scan# 1323  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050






Tgt Ion: 97 Resp: 21098


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 83.7  | 67.4  | 101.0 |
| 85  | 55.1  | 43.2  | 64.8  |
| 99  | 62.3  | 50.2  | 75.2  |

Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023






#56  
Ethyl methacrylate  
Concen: 52.388 ug/l  
RT: 9.116 min Scan# 1317  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43



Tgt Ion: 69 Resp: 351553

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 69  | 100   |       |       |
| 41  | 70.4  | 55.1  | 82.7  |
| 39  | 40.0  | 28.5  | 42.7  |





#57

1,3-Dichloropropane

Concen: 52.428 ug/l

RT: 9.305 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX034280.D

Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 36184

Ion Ratio Lower Upper

76 100





78 32.8 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 03/02/2023

Supervised By :Mahesh Dadoda 03/02/2023



#58

2-Chloroethyl Vinyl ether

Concen: 243.048 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX034280.D




Acq: 01 Mar 2023 09:43


Tgt Ion: 63 Resp: 706530

Ion Ratio Lower Upper

63 100

106 28.0 24.5 36.7





#59

2-Hexanone

Concen: 265.703 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX034280.D

Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 120251

Ion Ratio Lower Upper

43 100





58 53.2 27.4 82.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 03/02/2023

Supervised By :Mahesh Dadoda 03/02/2023



#60

Dibromochloromethane

Concen: 53.891 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX034280.D




Acq: 01 Mar 2023 09:43


Tgt Ion:129 Resp: 236209

Ion Ratio Lower Upper

129 100


127 77.2 38.8 116.3





#61  
1,2-Dibromoethane  
Concen: 53.879 ug/l  
RT: 9.610 min Scan# 1398  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:107 Resp: 226039  
Ion Ratio Lower Upper  
107 100  
109 93.6 74.2 111.2

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#62  
4-Bromofluorobenzene  
Concen: 55.516 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 95 Resp: 282150  
Ion Ratio Lower Upper  
95 100  
174 73.7 0.0 184.4  
176 70.7 0.0 178.8





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:117 Resp: 52588  
Ion Ratio Lower Upper  
117 100  
82 55.2 40.6 60.8  
119 32.9 25.6 38.4

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#64  
Tetrachloroethene  
Concen: 50.239 ug/l  
RT: 9.269 min Scan# 1342  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion:164 Resp: 186463  
Ion Ratio Lower Upper  
164 100  
166 127.3 100.2 150.2  
129 97.6 69.2 103.8  
131 93.5 67.2 100.8





#65  
Chlorobenzene  
Concen: 51.876 ug/l  
RT: 10.079 min Scan# 1475  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:112 Resp: 56866  
Ion Ratio Lower Upper  
112 100  
114 31.9 26.2 39.4

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#66  
1,1,1,2-Tetrachloroethane  
Concen: 50.571 ug/l  
RT: 10.159 min Scan# 1488  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43



Tgt Ion:131 Resp: 214605  
Ion Ratio Lower Upper  
131 100  
133 94.9 48.4 145.0  
119 66.2 31.9 95.7





#67

Ethyl Benzene

Concen: 52.740 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX034280.D

Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 1038654

Ion Ratio Lower Upper

91 100





106 31.3 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 03/02/2023

Supervised By :Mahesh Dadoda 03/02/2023



#68

m/p-Xylenes

Concen: 105.384 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX034280.D




Acq: 01 Mar 2023 09:43


Tgt Ion:106 Resp: 794229

Ion Ratio Lower Upper

106 100

91 205.1 156.4 234.6









#69  
o-Xylene  
Concen: 52.825 ug/l  
RT: 10.640 min Scan# 1567  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion:106 Resp: 39634  
Ion Ratio Lower Upper  
106 100  
91 214.8 103.1 309.3




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#70  
Styrene  
Concen: 53.667 ug/l  
RT: 10.652 min Scan# 1569  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43


Tgt Ion:104 Resp: 654886  
Ion Ratio Lower Upper  
104 100  
78 52.7 37.5 56.3  
103 54.7 43.3 64.9





#71  
Bromoform  
Concen: 51.138 ug/l  
RT: 10.799 min Scan# 1593  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:173 Resp: 17713  
Ion Ratio Lower Upper  
173 100  
175 48.3 24.1 72.3  
254 0.1 0.1 0.1

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43



Tgt Ion:152 Resp: 266228  
Ion Ratio Lower Upper  
152 100  
115 86.8 38.7 116.1  
150 174.0 0.0 348.2





#73  
Isopropylbenzene  
Concen: 50.575 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC050

Tgt Ion:105 Resp: 1026159  
Ion Ratio Lower Upper  
105 100  
120 27.2 13.6 40.8

Manual Integrations




APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#74  
N-ethyl acetate  
Concen: 48.377 ug/l  
RT: 10.841 min Scan# 1600  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Tgt Ion: 43 Resp: 493129  
Ion Ratio Lower Upper  
43 100  
70 40.4 34.8 52.2  
55 22.4 19.4 29.0  
61 23.5 19.4 29.2









#75  
1,1,2,2-Tetrachloroethane  
Concen: 47.459 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion: 83 Resp: 32935  
Ion Ratio Lower Upper  
83 100  
131 10.1 5.5 16.7  
85 64.3 32.3 96.9




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#76  
1,2,3-Trichloropropane  
Concen: 41.632 ug/l m  
RT: 11.238 min Scan# 1665  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43


Tgt Ion: 75 Resp: 278897  
Ion Ratio Lower Upper  
75 100  
77 44.4 21.8 65.4





#77  
Bromobenzene  
Concen: 49.977 ug/l  
RT: 11.195 min Scan# 1658  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:156 Resp: 244028  
Ion Ratio Lower Upper  
156 100  
77 171.4 68.1 204.3  
158 96.4 48.0 144.2

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#78  
n-propylbenzene  
Concen: 51.737 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43



Tgt Ion: 91 Resp: 1220864  
Ion Ratio Lower Upper  
91 100  
120 23.4 12.3 36.8





#79

2-Chlorotoluene

Concen: 50.944 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX034280.D

Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 73713

Ion Ratio Lower Upper

91 100





126 32.8 18.1 54.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 03/02/2023

Supervised By :Mahesh Dadoda 03/02/2023



#80

1,3,5-Trimethylbenzene

Concen: 50.686 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX034280.D




Acq: 01 Mar 2023 09:43


Tgt Ion:105 Resp: 864649

Ion Ratio Lower Upper

105 100


120 48.7 25.6 76.8





#81  
trans-1,4-Dichloro-2-butene  
Concen: 45.925 ug/l  
RT: 11.018 min Scan# 1629  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 75 Resp: 119759  
Ion Ratio Lower Upper  
75 100  
53 97.1 83.0 124.6  
89 44.7 37.7 56.5

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#82  
4-Chlorotoluene  
Concen: 51.053 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Tgt Ion: 91 Resp: 841379  
Ion Ratio Lower Upper  
91 100  
126 28.9 15.8 47.4





#83  
tert-Butylbenzene  
Concen: 49.353 ug/l  
RT: 11.713 min Scan# 1743  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:119 Resp: 86225  
Ion Ratio Lower Upper  
119 100  
91 60.9 26.9 80.7  
134 23.6 12.3 36.8

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023




#84  
1,2,4-Trimethylbenzene  
Concen: 50.746 ug/l  
RT: 11.750 min Scan# 1749  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43



Tgt Ion:105 Resp: 875272  
Ion Ratio Lower Upper  
105 100  
120 44.7 23.4 70.2





#85  
sec-Butylbenzene  
Concen: 51.503 ug/l  
RT: 11.890 min Scan# 1772  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050





Tgt Ion:105 Resp: 109272  
Ion Ratio Lower Upper  
105 100  
134 20.4 10.8 32.3

Manual Integrations

APPROVED




Reviewed By :Krupa Patel 03/02/2023


Supervised By :Mahesh Dadoda 03/02/2023



#86  
p-Isopropyltoluene  
Concen: 52.049 ug/l  
RT: 12.006 min Scan# 1791  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Tgt Ion:119 Resp: 913919  
Ion Ratio Lower Upper  
119 100  
134 26.6 13.6 40.8  
91 25.6 11.0 33.0





#87

1,3-Dichlorobenzene

Concen: 51.207 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX034280.D

Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050

Tgt Ion:146 Resp: 468918

Ion Ratio Lower Upper

146 100

111 41.9 18.6 55.6





148 63.7 31.9 95.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 03/02/2023

Supervised By :Mahesh Dadoda 03/02/2023



#88

1,4-Dichlorobenzene

Concen: 50.697 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX034280.D

Acq: 01 Mar 2023 09:43




Tgt Ion:146 Resp: 468257


Ion Ratio Lower Upper

146 100

111 41.0 17.8 53.4


148 63.4 32.1 96.3





#89  
n-Butylbenzene  
Concen: 53.069 ug/l  
RT: 12.329 min Scan# 1844  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050



Tgt Ion: 91 Resp: 829952  
Ion Ratio Lower Upper  
91 100  
92 54.7 27.5 82.5  
134 25.6 14.3 42.9




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#90  
Hexachloroethane  
Concen: 49.638 ug/l  
RT: 12.536 min Scan# 1878  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Tgt Ion:117 Resp: 182235  
Ion Ratio Lower Upper  
117 100  
201 70.6 45.1 135.3





#91  
1,2-Dichlorobenzene  
Concen: 50.474 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050





Tgt Ion:146 Resp: 45222  
Ion Ratio Lower Upper  
146 100  
111 43.2 19.3 57.8  
148 63.9 32.3 96.9

Manual Integrations

APPROVED




Reviewed By :Krupa Patel 03/02/2023


Supervised By :Mahesh Dadoda 03/02/2023



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 44.155 ug/l  
RT: 12.939 min Scan# 1944  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Tgt Ion: 75 Resp: 73071  
Ion Ratio Lower Upper  
75 100  
155 81.6 51.7 155.1  
157 105.8 67.5 202.3









#93  
1,2,4-Trichlorobenzene  
Concen: 50.654 ug/l  
RT: 13.585 min Scan# 2050  
Delta R.T. -0.006 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion:180 Resp: 29136  
Ion Ratio Lower Upper  
180 100  
182 94.8 47.9 143.6  
145 32.0 14.1 42.2




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#94  
Hexachlorobutadiene  
Concen: 51.588 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Tgt Ion:225 Resp: 123015  
Ion Ratio Lower Upper  
225 100  
223 62.5 31.3 93.9  
227 63.0 32.3 96.9





#95  
Naphthalene  
Concen: 47.984 ug/l  
RT: 13.774 min Scan# 2081  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC050

Tgt Ion:128 Resp: 929798  
Ion Ratio Lower Upper  
128 100  
127 12.8 10.3 15.5  
129 11.0 8.7 13.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 03/02/2023  
Supervised By :Mahesh Dadoda 03/02/2023



#96  
1,2,3-Trichlorobenzene  
Concen: 49.625 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. 0.000 min  
Lab File: VX034280.D  
Acq: 01 Mar 2023 09:43

Tgt Ion:180 Resp: 283392  
Ion Ratio Lower Upper  
180 100  
182 93.9 47.6 142.9  
145 33.9 15.1 45.3

