

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030124\
 Data File : VX040554.D
 Acq On : 01 Mar 2024 17:54
 Operator : JC/MD
 Sample : P1660-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-940-942

Quant Time: Mar 01 23:21:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70774	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	120238	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113612	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57644	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64783	54.727	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.460%	
35) Dibromofluoromethane	5.385	113	45867	54.856	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.720%	
50) Toluene-d8	8.647	98	164429	54.177	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.360%	
62) 4-Bromofluorobenzene	11.079	95	57016	49.282	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.560%	
Target Compounds						
16) Acetone	2.380	43	1646	4.264	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030124\
Data File : VX040554.D
Acq On : 01 Mar 2024 17:54
Operator : JC/MD
Sample : P1660-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-940-942

Quant Time: Mar 01 23:21:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

