

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030124\
 Data File : VX040556.D
 Acq On : 01 Mar 2024 18:40
 Operator : JC/MD
 Sample : P1660-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-EB-20240227

Quant Time: Mar 01 23:22:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72330	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	122166	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119898	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56920	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66721	55.152	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 110.300%			
35) Dibromofluoromethane	5.385	113	48008	56.511	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 113.020%			
50) Toluene-d8	8.647	98	170192	55.191	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 110.380%			
62) 4-Bromofluorobenzene	11.079	95	68431	58.215	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 116.440%			
Target Compounds						
16) Acetone	2.374	43	2242	5.683	ug/l	95
25) 2-Butanone	4.574	43	2314	3.366	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030124\
Data File : VX040556.D
Acq On : 01 Mar 2024 18:40
Operator : JC/MD
Sample : P1660-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-EB-20240227

Quant Time: Mar 01 23:22:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

