

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030124\
 Data File : VX040557.D
 Acq On : 01 Mar 2024 19:03
 Operator : JC/MD
 Sample : P1660-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-TB-20240227

Quant Time: Mar 01 23:22:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	68592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	118055	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	101263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57738	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62851	54.784	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 109.560%		
35) Dibromofluoromethane	5.385	113	44199	53.839	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 107.680%		
50) Toluene-d8	8.647	98	158875	53.315	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 106.620%		
62) 4-Bromofluorobenzene	11.079	95	63471	55.876	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 111.760%		
Target Compounds						
16) Acetone	2.380	43	954	2.550	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030124\
Data File : VX040557.D
Acq On : 01 Mar 2024 19:03
Operator : JC/MD
Sample : P1660-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-TB-20240227

Quant Time: Mar 01 23:22:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

