

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030124\
 Data File : VX040558.D
 Acq On : 01 Mar 2024 19:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 01 23:22:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	50.000	48.749	2.5	85	0.00
3 P	Chloromethane	50.000	49.756	0.5	88	0.00
4 C	Vinyl Chloride	50.000	50.962	-1.9#	90	0.00
5 T	Bromomethane	50.000	55.363	-10.7	86	0.00
6 T	Chloroethane	50.000	44.326	11.3	80	0.00
7 T	Trichlorofluoromethane	50.000	47.979	4.0	84	0.00
8 T	Diethyl Ether	50.000	47.161	5.7	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	52.032	-4.1	91	0.00
10 T	Methyl Iodide	50.000	53.779	-7.6	95	0.00
11 T	Tert butyl alcohol	250.000	302.035	-20.8	102	0.00
12 CM	1,1-Dichloroethene	50.000	47.578	4.8#	86	0.00
13 T	Acrolein	250.000	289.059	-15.6	103	0.00
14 T	Allyl chloride	50.000	50.968	-1.9	90	0.00
15 T	Acrylonitrile	250.000	268.545	-7.4	94	0.00
16 T	Acetone	250.000	323.589	-29.4#	110	0.00
17 T	Carbon Disulfide	50.000	42.397	15.2	77	0.00
18 T	Methyl Acetate	50.000	60.190	-20.4	101	0.00
19 T	Methyl tert-butyl Ether	50.000	54.519	-9.0	94	0.00
20 T	Methylene Chloride	50.000	50.933	-1.9	92	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.846	2.3	86	0.00
22 T	Diisopropyl ether	50.000	52.249	-4.5	90	0.00
23 T	Vinyl Acetate	250.000	268.842	-7.5	92	0.00
24 P	1,1-Dichloroethane	50.000	51.618	-3.2	91	0.00
25 T	2-Butanone	250.000	292.109	-16.8	99	0.00
26 T	2,2-Dichloropropane	50.000	47.797	4.4	85	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.910	-1.8	92	0.00
28 T	Bromochloromethane	50.000	52.372	-4.7	84	0.00
29 T	Tetrahydrofuran	250.000	272.074	-8.8	95	0.01
30 C	Chloroform	50.000	53.503	-7.0#	93	0.00
31 T	Cyclohexane	50.000	49.053	1.9	86	0.00
32 T	1,1,1-Trichloroethane	50.000	55.531	-11.1	97	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.297	1.4	89	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	84	0.00
35 S	Dibromofluoromethane	50.000	47.820	4.4	85	0.00
36 T	1,1-Dichloropropene	50.000	49.991	0.0	91	0.00
37 T	Ethyl Acetate	50.000	52.674	-5.3	93	0.00
38 T	Carbon Tetrachloride	50.000	54.584	-9.2	94	0.00
39 T	Methylcyclohexane	50.000	47.524	5.0	82	0.00
40 TM	Benzene	50.000	50.141	-0.3	88	0.00
41 T	Methacrylonitrile	50.000	53.866	-7.7	97	0.00
42 TM	1,2-Dichloroethane	50.000	56.892	-13.8	99	0.00
43 T	Isopropyl Acetate	50.000	54.653	-9.3	96	0.00
44 TM	Trichloroethene	50.000	51.067	-2.1	92	0.00
45 C	1,2-Dichloropropane	50.000	51.078	-2.2#	91	0.00
46 T	Dibromomethane	50.000	53.317	-6.6	93	0.00
47 T	Bromodichloromethane	50.000	55.164	-10.3	95	0.00
48 T	Methyl methacrylate	50.000	57.858	-15.7	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030124\
 Data File : VX040558.D
 Acq On : 01 Mar 2024 19:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 01 23:22:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1190.801	-19.1	99	0.00
50 S	Toluene-d8	50.000	46.262	7.5	81	0.00
51 T	4-Methyl-2-Pentanone	250.000	286.896	-14.8	98	0.00
52 CM	Toluene	50.000	52.360	-4.7#	91	0.00
53 T	t-1,3-Dichloropropene	50.000	54.932	-9.9	92	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.723	-3.4	90	0.00
55 T	1,1,2-Trichloroethane	50.000	55.581	-11.2	92	0.00
56 T	Ethyl methacrylate	50.000	56.742	-13.5	93	0.00
57 T	1,3-Dichloropropane	50.000	53.975	-8.0	93	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	248.706	0.5	81	0.00
59 T	2-Hexanone	250.000	295.159	-18.1	99	0.00
60 T	Dibromochloromethane	50.000	55.414	-10.8	95	0.00
61 T	1,2-Dibromoethane	50.000	55.149	-10.3	94	0.00
62 S	4-Bromofluorobenzene	50.000	43.595	12.8	75	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	77	0.00
64 T	Tetrachloroethene	50.000	58.869	-17.7	95	0.00
65 PM	Chlorobenzene	50.000	51.008	-2.0	83	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	55.438	-10.9	87	0.00
67 C	Ethyl Benzene	50.000	52.380	-4.8#	82	0.00
68 T	m/p-Xylenes	100.000	104.871	-4.9	83	0.00
69 T	o-Xylene	50.000	53.309	-6.6	83	0.00
70 T	Styrene	50.000	53.971	-7.9	84	0.00
71 P	Bromoform	50.000	56.817	-13.6	87	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	79	0.00
73 T	Isopropylbenzene	50.000	51.466	-2.9	83	0.00
74 T	N-ethyl acetate	50.000	54.195	-8.4	85	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.594	-3.2	85	0.00
76 T	1,2,3-Trichloropropane	50.000	52.997	-6.0	84	0.00
77 T	Bromobenzene	50.000	52.401	-4.8	84	0.00
78 T	n-propylbenzene	50.000	50.403	-0.8	81	0.00
79 T	2-Chlorotoluene	50.000	49.636	0.7	82	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.388	-4.8	84	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.119	5.8	79	0.00
82 T	4-Chlorotoluene	50.000	51.280	-2.6	83	0.00
83 T	tert-Butylbenzene	50.000	53.191	-6.4	85	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.472	-4.9	83	0.00
85 T	sec-Butylbenzene	50.000	52.848	-5.7	84	0.00
86 T	p-Isopropyltoluene	50.000	52.285	-4.6	84	0.00
87 T	1,3-Dichlorobenzene	50.000	50.246	-0.5	83	0.00
88 T	1,4-Dichlorobenzene	50.000	50.321	-0.6	85	0.00
89 T	n-Butylbenzene	50.000	50.503	-1.0	81	0.00
90 T	Hexachloroethane	50.000	54.072	-8.1	86	0.00
91 T	1,2-Dichlorobenzene	50.000	51.710	-3.4	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	57.453	-14.9	90	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.687	-7.4	86	0.00
94 T	Hexachlorobutadiene	50.000	50.815	-1.6	85	0.00
95 T	Naphthalene	50.000	54.937	-9.9	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030124\
Data File : VX040558.D
Acq On : 01 Mar 2024 19:25
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 01 23:22:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	54.952	-9.9	88	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6