

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030220\
 Data File : VX015062.D
 Acq On : 02 Mar 2020 10:16
 Operator : JC/SP
 Sample : VX0302WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0302WBL01

Quant Time: Mar 02 16:30:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	437739	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	693390	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	644210	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	308973	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	248774	50.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.72%	
35) Dibromofluoromethane	5.48	113	214087	50.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.44%	
50) Toluene-d8	8.71	98	832102	51.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.28%	
62) 4-Bromofluorobenzene	11.13	95	275864	47.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.12%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030220\
Data File : VX015062.D
Acq On : 02 Mar 2020 10:16
Operator : JC/SP
Sample : VX0302WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0302WBL01

Quant Time: Mar 02 16:30:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

