

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030220\
 Data File : VX015066.D
 Acq On : 02 Mar 2020 12:22
 Operator : JC/SP
 Sample : L1724-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A2828

Manual Integrations
 APPROVED

MMDadoda
 3/3/2020 9:51:35 AM

Quant Time: Mar 02 16:48:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	412239	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	659146	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	624422	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	329832	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	238918	51.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.70%	
35) Dibromofluoromethane	5.48	113	205737	50.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.54%	
50) Toluene-d8	8.71	98	747742	48.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.68%	
62) 4-Bromofluorobenzene	11.13	95	308343	55.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.68%	
Target Compounds						
					Qvalue	
16) Acetone	2.43	43	36675	3.918	ug/l	98
30) Chloroform	5.20	83	25241	3.523	ug/l	87
40) Benzene	6.14	78	6651	0.378	ug/l #	87
47) Bromodichloromethane	7.89	83	5925	0.973	ug/l #	95
52) Toluene	8.78	92	11105	0.992	ug/l	90
67) Ethyl Benzene	10.25	91	21854	1.015	ug/l	99
68) m/p-Xylenes	10.35	106	7014	0.845	ug/l	96
69) o-Xylene	10.69	106	14964	1.876	ug/l	94
73) Isopropylbenzene	11.01	105	11758	0.568	ug/l	86
78) n-propylbenzene	11.35	91	16782	0.701	ug/l	99
80) 1,3,5-Trimethylbenzene	11.50	105	28549	1.652	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	44560	2.558	ug/l	96
85) sec-Butylbenzene	11.94	105	8496m	0.418	ug/l	
86) p-Isopropyltoluene	12.06	119	82760	4.350	ug/l #	77
89) n-Butylbenzene	12.38	91	23201	1.362	ug/l #	52

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030220\
Data File : VX015066.D
Acq On : 02 Mar 2020 12:22
Operator : JC/SP
Sample : L1724-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A2828

Manual Integrations
APPROVED

MMDadoda
3/3/2020 9:51:35 AM

Quant Time: Mar 02 16:48:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

