

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
 Data File : VX020942.D
 Acq On : 01 Mar 2021 18:23
 Operator : JC/MD
 Sample : PB134860ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB134860ZHE#01

Quant Time: Mar 02 00:31:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	198749	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	360462	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	338321	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	143769	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	126949	47.29	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.58%
35) Dibromofluoromethane	5.470	113	110633	50.52	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.04%
50) Toluene-d8	8.694	98	432061	50.60	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.20%
62) 4-Bromofluorobenzene	11.118	95	161889	51.27	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.54%
Target Compounds						
					Qvalue	
16) Acetone	2.417	43	38733	35.01	ug/l	100
20) Methylene Chloride	2.834	84	30433	11.47	ug/l	97
43) Isopropyl Acetate	6.405	43	19189	3.46	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
Data File : VX020942.D
Acq On : 01 Mar 2021 18:23
Operator : JC/MD
Sample : PB134860ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB134860ZHE#01

Quant Time: Mar 02 00:31:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

