

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
 Data File : VX020943.D
 Acq On : 01 Mar 2021 18:46
 Operator : JC/MD
 Sample : PB134860ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB134860ZHE#02

Quant Time: Mar 02 00:31:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	200900	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	363836	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	355747	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	148248	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	127901	47.13	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.26%
35) Dibromofluoromethane	5.470	113	111391	50.39	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.78%
50) Toluene-d8	8.694	98	447377	51.90	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.80%
62) 4-Bromofluorobenzene	11.118	95	166577	52.27	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.54%
Target Compounds						
16) Acetone	2.416	43	38794	34.69	ug/l	99
20) Methylene Chloride	2.834	84	29848	11.13	ug/l	95
43) Isopropyl Acetate	6.405	43	19776	3.53	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
Data File : VX020943.D
Acq On : 01 Mar 2021 18:46
Operator : JC/MD
Sample : PB134860ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB134860ZHE#02

Quant Time: Mar 02 00:31:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

