

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
 Data File : VX020944.D
 Acq On : 01 Mar 2021 19:09
 Operator : JC/MD
 Sample : PB134860ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB134860ZHE#03

Quant Time: Mar 02 00:31:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	211109	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.829	114	376129	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	356634	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	147308	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	132141	46.34	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.68%
35) Dibromofluoromethane	5.458	113	117847	51.57	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.14%
50) Toluene-d8	8.694	98	459724	51.59	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.18%
62) 4-Bromofluorobenzene	11.118	95	168647	51.19	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.38%
Target Compounds						
					Qvalue	
16) Acetone	2.417	43	38722	32.96	ug/l	98
20) Methylene Chloride	2.834	84	30295	10.75	ug/l	97
43) Isopropyl Acetate	6.405	43	19757	3.42	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
Data File : VX020944.D
Acq On : 01 Mar 2021 19:09
Operator : JC/MD
Sample : PB134860ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB134860ZHE#03

Quant Time: Mar 02 00:31:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

