

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
 Data File : VX020950.D
 Acq On : 01 Mar 2021 21:28
 Operator : JC/MD
 Sample : M1522-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-GW-718-720

Quant Time: Mar 02 00:33:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	188062	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	339018	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	330050	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	149614	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	119895	47.20	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 94.40%			
35) Dibromofluoromethane	5.464	113	105721	51.33	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.66%			
50) Toluene-d8	8.694	98	404930	50.42	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 100.84%			
62) 4-Bromofluorobenzene	11.118	95	160099	53.91	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 107.82%			
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.364	101	722	0.36	ug/l	97
30) Chloroform	5.176	83	1309	0.32	ug/l #	77
38) Carbon Tetrachloride	5.764	117	1302	0.42	ug/l #	77
44) Trichloroethene	7.188	130	13441	5.29	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
Data File : VX020950.D
Acq On : 01 Mar 2021 21:28
Operator : JC/MD
Sample : M1522-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW6-GW-718-720

Quant Time: Mar 02 00:33:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

