

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
 Data File : VX020953.D
 Acq On : 01 Mar 2021 22:37
 Operator : JC/MD
 Sample : M1522-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-GW-688-690

Quant Time: Mar 02 00:34:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	197799	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	356404	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	349361	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	146963	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	128860	48.23	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.46%
35) Dibromofluoromethane	5.464	113	112669	52.03	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.06%
50) Toluene-d8	8.694	98	443562	52.53	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.06%
62) 4-Bromofluorobenzene	11.118	95	158089	50.64	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.28%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.187	85	2242	1.05	ug/l 96
9) 1,1,2-Trichlorotrifluo...	2.370	101	12312	5.79	ug/l 97
12) 1,1-Dichloroethene	2.358	96	3090	1.40	ug/l 89
16) Acetone	2.417	43	2868	2.61	ug/l 93
27) cis-1,2-Dichloroethene	4.564	96	1927	0.67	ug/l 96
30) Chloroform	5.176	83	4184	0.97	ug/l 89
32) 1,1,1-Trichloroethane	5.458	97	885	0.24	ug/l # 49
38) Carbon Tetrachloride	5.752	117	6060	1.85	ug/l 99
44) Trichloroethene	7.188	130	58926	22.07	ug/l 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
Data File : VX020953.D
Acq On : 01 Mar 2021 22:37
Operator : JC/MD
Sample : M1522-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW6-GW-688-690

Quant Time: Mar 02 00:34:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

