

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
 Data File : VX020954.D
 Acq On : 01 Mar 2021 23:00
 Operator : JC/MD
 Sample : M1522-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-DUP-20210224

Quant Time: Mar 02 00:34:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	199353	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.829	114	365666	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	346887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	141920	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	130354	48.41	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.82%
35) Dibromofluoromethane	5.464	113	111858	50.35	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.70%
50) Toluene-d8	8.694	98	432103	49.88	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.76%
62) 4-Bromofluorobenzene	11.118	95	161600	50.45	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.90%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.187	85	2314	1.07	ug/l	91
9) 1,1,2-Trichlorotrifluo...	2.370	101	11952	5.57	ug/l	99
12) 1,1-Dichloroethene	2.358	96	2978	1.33	ug/l	93
16) Acetone	2.423	43	2011	1.81	ug/l	99
27) cis-1,2-Dichloroethene	4.564	96	1764	0.61	ug/l	91
30) Chloroform	5.176	83	4204	0.97	ug/l	85
32) 1,1,1-Trichloroethane	5.458	97	996	0.26	ug/l #	50
38) Carbon Tetrachloride	5.752	117	6068	1.80	ug/l	94
44) Trichloroethene	7.188	130	59545	21.74	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
Data File : VX020954.D
Acq On : 01 Mar 2021 23:00
Operator : JC/MD
Sample : M1522-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW6-DUP-20210224

Quant Time: Mar 02 00:34:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

