

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
 Data File : VX020960.D
 Acq On : 02 Mar 2021 01:19
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 02 01:45:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.541	0.521	3.7	97	0.00
3 P	Chloromethane	0.601	0.606	-0.8	102	0.00
4 C	Vinyl Chloride	0.692	0.683	1.3#	100	0.00
5 T	Bromomethane	0.249	0.245	1.6	114	0.00
6 T	Chloroethane	0.433	0.449	-3.7	105	0.00
7 T	Trichlorofluoromethane	0.873	0.890	-1.9	105	0.00
8 T	Diethyl Ether	0.405	0.407	-0.5	105	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.538	0.524	2.6	99	0.00
10 T	Methyl Iodide	0.723	0.795	-10.0	105	0.00
11 T	Tert butyl alcohol	0.150	0.131	12.7	87	0.00
12 CM	1,1-Dichloroethene	0.560	0.556	0.7#	101	0.00
13 T	Acrolein	0.093	0.101	-8.6	114	0.00
14 T	Allyl chloride	0.914	0.859	6.0	95	0.00
15 T	Acrylonitrile	0.321	0.313	2.5	98	0.00
16 T	Acetone	0.278	0.245	11.9	94	0.00
17 T	Carbon Disulfide	1.594	1.569	1.6	100	0.00
18 T	Methyl Acetate	0.658	0.614	6.7	94	0.00
19 T	Methyl tert-butyl Ether	1.999	1.975	1.2	99	0.00
20 T	Methylene Chloride	0.668	0.659	1.3	105	0.00
21 T	trans-1,2-Dichloroethene	0.629	0.630	-0.2	103	0.00
22 T	Diisopropyl ether	1.839	1.821	1.0	101	0.00
23 T	Vinyl Acetate	1.572	1.540	2.0	98	0.00
24 P	1,1-Dichloroethane	1.083	1.074	0.8	101	0.00
25 T	2-Butanone	0.423	0.393	7.1	95	0.00
26 T	2,2-Dichloropropane	0.984	0.825	16.2	86	0.00
27 T	cis-1,2-Dichloroethene	0.724	0.732	-1.1	104	0.00
28 T	Bromochloromethane	0.463	0.456	1.5	101	0.00
29 T	Tetrahydrofuran	0.271	0.254	6.3	96	0.00
30 C	Chloroform	1.091	1.140	-4.5#	106	0.00
31 T	Cyclohexane	0.978	0.944	3.5	98	0.00
32 T	1,1,1-Trichloroethane	0.950	0.958	-0.8	103	0.00
33 S	1,2-Dichloroethane-d4	0.675	0.667	1.2	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.304	0.312	-2.6	103	0.00
36 T	1,1-Dichloropropene	0.475	0.483	-1.7	102	0.00
37 T	Ethyl Acetate	0.452	0.442	2.2	96	0.00
38 T	Carbon Tetrachloride	0.460	0.462	-0.4	104	0.00
39 T	Methylcyclohexane	0.596	0.570	4.4	98	0.00
40 TM	Benzene	1.424	1.471	-3.3	104	0.00
41 T	Methacrylonitrile	0.254	0.250	1.6	98	0.00
42 TM	1,2-Dichloroethane	0.474	0.494	-4.2	104	0.00
43 T	Isopropyl Acetate	0.769	0.751	2.3	97	0.00
44 TM	Trichloroethene	0.375	0.400	-6.7	107	0.00
45 C	1,2-Dichloropropane	0.357	0.367	-2.8#	103	0.00
46 T	Dibromomethane	0.240	0.257	-7.1	107	0.00
47 T	Bromodichloromethane	0.501	0.528	-5.4	106	0.00
48 T	Methyl methacrylate	0.386	0.381	1.3	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
 Data File : VX020960.D
 Acq On : 02 Mar 2021 01:19
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 02 01:45:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	96	0.00
50 S	Toluene-d8	1.185	1.190	-0.4	102	0.00
51 T	4-Methyl-2-Pentanone	0.448	0.442	1.3	97	0.00
52 CM	Toluene	0.887	0.953	-7.4#	108	0.00
53 T	t-1,3-Dichloropropene	0.561	0.569	-1.4	99	0.00
54 T	cis-1,3-Dichloropropene	0.615	0.618	-0.5	100	0.00
55 T	1,1,2-Trichloroethane	0.351	0.377	-7.4	107	0.00
56 T	Ethyl methacrylate	0.567	0.574	-1.2	100	0.00
57 T	1,3-Dichloropropane	0.601	0.630	-4.8	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.277	0.286	-3.2	99	0.00
59 T	2-Hexanone	0.336	0.326	3.0	95	0.00
60 T	Dibromochloromethane	0.370	0.397	-7.3	103	0.00
61 T	1,2-Dibromoethane	0.367	0.402	-9.5	109	0.00
62 S	4-Bromofluorobenzene	0.438	0.440	-0.5	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.361	0.369	-2.2	109	0.00
65 PM	Chlorobenzene	1.053	1.066	-1.2	106	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.396	-2.1	105	0.00
67 C	Ethyl Benzene	1.909	1.932	-1.2#	107	0.00
68 T	m/p-Xylenes	0.703	0.716	-1.8	108	0.00
69 T	o-Xylene	0.686	0.689	-0.4	104	0.00
70 T	Styrene	1.176	1.198	-1.9	106	0.00
71 P	Bromoform	0.300	0.299	0.3	106	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.949	4.098	-3.8	109	0.00
74 T	N-amyl acetate	1.599	1.498	6.3	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.273	1.251	1.7	101	0.00
76 T	1,2,3-Trichloropropane	1.076	1.061	1.4	100	0.00
77 T	Bromobenzene	0.919	1.008	-9.7	112	0.00
78 T	n-propylbenzene	4.504	4.609	-2.3	105	0.00
79 T	2-Chlorotoluene	2.729	2.725	0.1	104	0.00
80 T	1,3,5-Trimethylbenzene	3.286	3.331	-1.4	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.486	0.422	13.2	88	0.00
82 T	4-Chlorotoluene	3.145	3.172	-0.9	103	0.00
83 T	tert-Butylbenzene	3.197	3.202	-0.2	105	0.00
84 T	1,2,4-Trimethylbenzene	3.336	3.354	-0.5	102	0.00
85 T	sec-Butylbenzene	3.768	3.808	-1.1	102	0.00
86 T	p-Isopropyltoluene	3.453	3.433	0.6	101	0.00
87 T	1,3-Dichlorobenzene	1.645	1.682	-2.2	106	0.00
88 T	1,4-Dichlorobenzene	1.676	1.760	-5.0	109	0.00
89 T	n-Butylbenzene	3.083	3.012	2.3	100	0.00
90 T	Hexachloroethane	0.659	0.631	4.2	97	0.00
91 T	1,2-Dichlorobenzene	1.606	1.563	2.7	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.292	0.268	8.2	94	0.00
93 T	1,2,4-Trichlorobenzene	1.030	1.034	-0.4	100	0.00
94 T	Hexachlorobutadiene	0.413	0.403	2.4	101	0.00
95 T	Naphthalene	3.667	3.865	-5.4	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030221\
Data File : VX020960.D
Acq On : 02 Mar 2021 01:19
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 02 01:45:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.023	1.102	-7.7	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6