

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030222\
 Data File : VX027263.D
 Acq On : 02 Mar 2022 12:39
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 03 00:04:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72469	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	116205	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	107169	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	42325	48.629	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.260%
35) Dibromofluoromethane	5.391	113	35715	49.515	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.020%
50) Toluene-d8	8.653	98	131925	48.915	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.820%
62) 4-Bromofluorobenzene	11.079	95	52265	50.485	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.960%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	32285	44.818	ug/l	99
3) Chloromethane	1.294	50	27452	40.768	ug/l	100
4) Vinyl Chloride	1.374	62	27212	40.571	ug/l	97
5) Bromomethane	1.599	94	12676	43.147	ug/l	99
6) Chloroethane	1.678	64	17485	42.453	ug/l	99
7) Trichlorofluoromethane	1.886	101	48876	46.853	ug/l	100
8) Diethyl Ether	2.136	74	15974	43.994	ug/l	89
9) 1,1,2-Trichlorotrifluo...	2.331	101	35178	48.406	ug/l	96
10) Methyl Iodide	2.453	142	29866	30.606	ug/l	95
11) Tert butyl alcohol	2.971	59	40921	241.647	ug/l	96
12) 1,1-Dichloroethene	2.318	96	33763	47.856	ug/l	87
13) Acrolein	2.239	56	39725	253.796	ug/l	98
14) Allyl chloride	2.666	41	58885	47.405	ug/l	95
15) Acrylonitrile	3.068	53	108753	250.273	ug/l	98
16) Acetone	2.386	43	86830	227.222	ug/l	95
17) Carbon Disulfide	2.514	76	82985	45.236	ug/l	99
18) Methyl Acetate	2.709	43	47076	47.728	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	119350	49.301	ug/l	100
20) Methylene Chloride	2.794	84	38203	48.233	ug/l	92
21) trans-1,2-Dichloroethene	3.093	96	37016	49.198	ug/l	92
22) Diisopropyl ether	3.763	45	117142	48.374	ug/l	98
23) Vinyl Acetate	3.727	43	512731	250.737	ug/l	96
24) 1,1-Dichloroethane	3.611	63	66960	48.914	ug/l	98
25) 2-Butanone	4.556	43	148057	243.287	ug/l	95
26) 2,2-Dichloropropane	4.483	77	49589	49.934	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	44163	50.613	ug/l	93
28) Bromochloromethane	4.897	49	26955	48.672	ug/l	87
29) Tetrahydrofuran	5.007	42	94254	231.407	ug/l	95
30) Chloroform	5.098	83	69031	48.550	ug/l	100
31) Cyclohexane	5.470	56	59980	46.015	ug/l	96
32) 1,1,1-Trichloroethane	5.391	97	61309	49.810	ug/l	99
36) 1,1-Dichloropropene	5.696	75	50547	48.017	ug/l	97
37) Ethyl Acetate	4.714	43	55426	47.955	ug/l	97
38) Carbon Tetrachloride	5.678	117	54722	49.841	ug/l	98
39) Methylcyclohexane	7.379	83	62787	47.948	ug/l	98
40) Benzene	6.037	78	149269	48.488	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030222\
 Data File : VX027263.D
 Acq On : 02 Mar 2022 12:39
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 03 00:04:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	30361	47.885	ug/l	95
42) 1,2-Dichloroethane	6.092	62	54354	49.770	ug/l	98
43) Isopropyl Acetate	6.342	43	87095	48.984	ug/l	96
44) Trichloroethene	7.129	130	43622	47.369	ug/l	96
45) 1,2-Dichloropropane	7.433	63	37161	49.405	ug/l	98
46) Dibromomethane	7.580	93	26968	49.917	ug/l	93
47) Bromodichloromethane	7.824	83	53048	50.307	ug/l	100
48) Methyl methacrylate	7.696	41	42635	49.539	ug/l	94
49) 1,4-Dioxane	7.659	88	19285	991.189	ug/l	94
51) 4-Methyl-2-Pentanone	8.573	43	293834	254.260	ug/l	98
52) Toluene	8.720	92	101565	51.524	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	54722	46.107	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	60936	51.769	ug/l	94
55) 1,1,2-Trichloroethane	9.153	97	40326	51.650	ug/l	98
56) Ethyl methacrylate	9.116	69	62013	51.016	ug/l	93
57) 1,3-Dichloropropane	9.311	76	65961	50.793	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	151300	258.226	ug/l	95
59) 2-Hexanone	9.433	43	219478	248.487	ug/l	96
60) Dibromochloromethane	9.525	129	43159	53.287	ug/l	99
61) 1,2-Dibromoethane	9.610	107	42520	51.138	ug/l	99
64) Tetrachloroethene	9.275	164	44401	45.793	ug/l	96
65) Chlorobenzene	10.079	112	111489	51.034	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.165	131	41020	51.778	ug/l	98
67) Ethyl Benzene	10.195	91	197668	51.543	ug/l	98
68) m/p-Xylenes	10.305	106	149041	101.090	ug/l	97
69) o-Xylene	10.646	106	74265	50.189	ug/l	95
70) Styrene	10.658	104	125940	52.014	ug/l	97
71) Bromoform	10.799	173	31867	46.844	ug/l #	98
73) Isopropylbenzene	10.963	105	194160	51.045	ug/l	98
74) N-amyl acetate	10.841	43	75328	52.049	ug/l	95
75) 1,1,2,2-Tetrachloroethane	11.213	83	57146	50.865	ug/l	98
76) 1,2,3-Trichloropropane	11.244	75	66487	62.096	ug/l	90
77) Bromobenzene	11.201	156	47862	50.447	ug/l	94
78) n-propylbenzene	11.305	91	218745	50.598	ug/l	98
79) 2-Chlorotoluene	11.366	91	134659	50.830	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	164420	51.407	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	16360	45.220	ug/l	91
82) 4-Chlorotoluene	11.457	91	153278	50.647	ug/l	96
83) tert-Butylbenzene	11.713	119	161439	51.909	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	167041	52.505	ug/l	98
85) sec-Butylbenzene	11.890	105	209744	53.786	ug/l	98
86) p-Isopropyltoluene	12.012	119	180168	54.195	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	91377	51.184	ug/l	98
88) 1,4-Dichlorobenzene	12.042	146	90741	49.711	ug/l	98
89) n-Butylbenzene	12.335	91	145191	52.253	ug/l	99
90) Hexachloroethane	12.542	117	27197	52.814	ug/l	93
91) 1,2-Dichlorobenzene	12.335	146	88256	50.666	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	13461	51.608	ug/l	87
93) 1,2,4-Trichlorobenzene	13.591	180	56850	51.719	ug/l	98
94) Hexachlorobutadiene	13.725	225	23484	51.460	ug/l	98
95) Naphthalene	13.780	128	195049	52.219	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	58174	52.799	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030222\
Data File : VX027263.D
Acq On : 02 Mar 2022 12:39
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Mar 03 00:04:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion:168 Resp: 72469
Ion Ratio Lower Upper
168 100
99 56.1 50.3 75.5




#2
Dichlorodifluoromethane
Concen: 44.818 ug/l
RT: 1.172 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion: 85 Resp: 32285
Ion Ratio Lower Upper
85 100
87 32.0 16.4 49.1



Abundance Scan 34 (1.294 min): VX027138.D\data.ms (-29)



#3

Chloromethane

Concen: 40.768 ug/l

RT: 1.294 min Scan# 34

Delta R.T. 0.000 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X

ClientSampleId :


Tgt Ion: 50 Resp: 27452

Ion Ratio Lower Upper


50 100

52 32.8 26.1 39.1


Abundance Scan 34 (1.294 min): VX027263.D\data.ms




Abundance



Abundance Scan 34 (1.294 min): VX027263.D\data.ms (-1) (



Abundance Scan 48 (1.380 min): VX027138.D\data.ms (-43)



#4

Vinyl Chloride

Concen: 40.571 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.006 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39


Tgt Ion: 62 Resp: 27212

Ion Ratio Lower Upper


62 100

64 31.1 26.2 39.2


Abundance Scan 47 (1.374 min): VX027263.D\data.ms






Abundance



Abundance Scan 47 (1.374 min): VX027263.D\data.ms (-1) (



Abundance Scan 84 (1.599 min): VX027138.D\data.ms (-80)



#5

Bromomethane

Concen: 43.147 ug/l

RT: 1.599 min Scan# 84

Delta R.T. 0.000 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X


ClientSampleId :

Tgt Ion: 94 Resp: 12676




Ion Ratio Lower Upper

94 100

96 93.8 75.9 113.9



Abundance Scan 97 (1.678 min): VX027138.D\data.ms (-91)



#6

Chloroethane

Concen: 42.453 ug/l

RT: 1.678 min Scan# 97

Delta R.T. 0.000 min

Lab File: VX027263.D


Acq: 02 Mar 2022 12:39


Tgt Ion: 64 Resp: 17485

Ion Ratio Lower Upper

64 100

66 31.0 25.4 38.2









#7
 Trichlorofluoromethane
 Concen: 46.853 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX027263.D
 Acq: 02 Mar 2022 12:39




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:101 Resp: 48876
 Ion Ratio Lower Upper
 101 100
 103 64.1 51.3 76.9



#8
 Diethyl Ether
 Concen: 43.994 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX027263.D
 Acq: 02 Mar 2022 12:39

Tgt Ion: 74 Resp: 15974
 Ion Ratio Lower Upper
 74 100
 45 90.0 50.7 152.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.406 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 35178
Ion Ratio Lower Upper
101 100
85 42.3 35.3 52.9
151 81.2 61.4 92.2



#10
Methyl Iodide
Concen: 30.606 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion:142 Resp: 29866
Ion Ratio Lower Upper
142 100
127 43.2 38.1 57.1
141 14.4 11.2 16.8









#11
Tert butyl alcohol
Concen: 241.647 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 59 Resp: 40921
Ion Ratio Lower Upper
59 100
57 10.6 7.2 10.8



#12
1,1-Dichloroethene
Concen: 47.856 ug/l
RT: 2.318 min Scan# 202
Delta R.T. -0.001 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion: 96 Resp: 33763
Ion Ratio Lower Upper
96 100
61 149.0 138.5 207.7
98 62.8 51.8 77.6









#13
Acrolein
Concen: 253.796 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 56 Resp: 39725
Ion Ratio Lower Upper
56 100
55 69.9 56.9 85.3



#14
Allyl chloride
Concen: 47.405 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion: 41 Resp: 58885
Ion Ratio Lower Upper
41 100
39 63.3 54.2 81.4
76 33.8 25.2 37.8









#15
Acrylonitrile
Concen: 250.273 ug/l
RT: 3.068 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 53 Resp: 108753
Ion Ratio Lower Upper
53 100
52 81.4 66.6 100.0
51 35.4 29.5 44.3



#16
Acetone
Concen: 227.222 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion: 43 Resp: 86830
Ion Ratio Lower Upper
43 100
58 31.6 23.0 34.6









#17
Carbon Disulfide
Concen: 45.236 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 82985
Ion Ratio Lower Upper
76 100
78 9.1 7.5 11.3



#18
Methyl Acetate
Concen: 47.728 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion: 43 Resp: 47076
Ion Ratio Lower Upper
43 100
74 24.8 18.3 27.5









#19
Methyl tert-butyl Ether
Concen: 49.301 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 119350
Ion Ratio Lower Upper
73 100
57 21.9 17.4 26.2



#20
Methylene Chloride
Concen: 48.233 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39


Tgt Ion: 84 Resp: 38203
Ion Ratio Lower Upper
84 100
49 114.0 101.5 152.3
51 34.3 31.2 46.8
86 65.7 51.3 76.9








#21
trans-1,2-Dichloroethene
Concen: 49.198 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 37016
Ion Ratio Lower Upper
96 100
61 129.0 112.9 169.3
98 64.0 49.5 74.3




#22
Diisopropyl ether
Concen: 48.374 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39



Tgt Ion: 45 Resp: 117142
Ion Ratio Lower Upper
45 100
43 77.3 60.6 91.0
87 30.2 22.0 33.0
59 11.6 9.0 13.4









#23
 Vinyl Acetate
 Concen: 250.737 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX027263.D
 Acq: 02 Mar 2022 12:39




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 512731
 Ion Ratio Lower Upper
 43 100
 86 11.6 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 48.914 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.006 min
 Lab File: VX027263.D
 Acq: 02 Mar 2022 12:39

Tgt Ion: 63 Resp: 66960
 Ion Ratio Lower Upper
 63 100
 98 8.4 3.7 11.1
 100 4.4 2.4 7.2





#25

2-Butanone

Concen: 243.287 ug/l

RT: 4.556 min Scan# 5

Delta R.T. -0.006 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 148057

Ion Ratio Lower Upper

43 100

72 27.8 20.4 30.6



#26

2,2-Dichloropropane

Concen: 49.934 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX027263.D




Acq: 02 Mar 2022 12:39

Tgt Ion: 77 Resp: 49589

Ion Ratio Lower Upper

77 100


97 23.7 11.0 33.0








#27
cis-1,2-Dichloroethene
Concen: 50.613 ug/l
RT: 4.489 min Scan# 51
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 44163
Ion Ratio Lower Upper
96 100
61 136.2 0.0 296.6
98 63.8 0.0 128.8




#28
Bromochloromethane
Concen: 48.672 ug/l
RT: 4.897 min Scan# 625
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39



Tgt Ion: 49 Resp: 26955
Ion Ratio Lower Upper
49 100
129 2.1 0.0 4.2
130 89.3 62.4 93.6









#29
Tetrahydrofuran
Concen: 231.407 ug/l
RT: 5.007 min Scan# 643
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 42 Resp: 94254
Ion Ratio Lower Upper
42 100
72 46.4 34.6 52.0
71 42.7 31.8 47.8



#30
Chloroform
Concen: 48.550 ug/l
RT: 5.098 min Scan# 658
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39


Tgt Ion: 83 Resp: 69031
Ion Ratio Lower Upper
83 100
85 64.4 51.6 77.4








#31
Cyclohexane
Concen: 46.015 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 59980
Ion Ratio Lower Upper
56 100
69 33.8 25.8 38.6
84 89.8 68.9 103.3



#32
1,1,1-Trichloroethane
Concen: 49.810 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion: 97 Resp: 61309
Ion Ratio Lower Upper
97 100
99 64.1 51.3 76.9
61 41.8 35.0 52.6





#33

1,2-Dichloroethane-d4

Concen: 48.629 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 42325

Ion Ratio Lower Upper

65 100

67 51.2 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39




Tgt Ion: 114 Resp: 116205

Ion Ratio Lower Upper

114 100

63 19.6 0.0 42.8


88 15.7 0.0 33.0








#35
Dibromofluoromethane
Concen: 49.515 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:113 Resp: 35715
Ion Ratio Lower Upper
113 100
111 103.9 81.2 121.8
192 20.5 15.0 22.4




#36
1,1-Dichloropropene
Concen: 48.017 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39



Tgt Ion: 75 Resp: 50547
Ion Ratio Lower Upper
75 100
110 37.6 17.2 51.6
77 31.0 24.6 36.8





#37

Ethyl Acetate

Concen: 47.955 ug/l

RT: 4.714 min Scan# 596

Delta R.T. -0.007 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 43 Resp: 55426

Ion Ratio Lower Upper

43 100

61 14.2 10.6 16.0

70 11.8 8.5 12.7



#38

Carbon Tetrachloride

Concen: 49.841 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.006 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39




Tgt Ion: 117 Resp: 54722

Ion Ratio Lower Upper

117 100

119 95.8 74.7 112.1

121 30.6 25.4 38.0





#39

Methylcyclohexane

Concen: 47.948 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. -0.006 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 62787

Ion Ratio Lower Upper

83 100

55 77.1 63.6 95.4

98 48.6 38.4 57.6



#40

Benzene

Concen: 48.488 ug/l

RT: 6.037 min Scan# 812

Delta R.T. -0.006 min

Lab File: VX027263.D




Acq: 02 Mar 2022 12:39


Tgt Ion: 78 Resp: 149269

Ion Ratio Lower Upper

78 100


77 24.3 18.9 28.3








#41
Methacrylonitrile
Concen: 47.885 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41 Resp: 30361
Ion Ratio Lower Upper
41 100
39 55.4 45.4 68.0
67 74.4 53.2 79.8
52 29.6 23.0 34.6



#42
1,2-Dichloroethane
Concen: 49.770 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39


Tgt Ion: 62 Resp: 54354
Ion Ratio Lower Upper
62 100
98 9.5 0.0 17.8








#43
Isopropyl Acetate
Concen: 48.984 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 87095
Ion Ratio Lower Upper
43 100
61 21.0 15.4 23.0
87 14.6 10.2 15.2



#44
Trichloroethene
Concen: 47.369 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion:130 Resp: 43622
Ion Ratio Lower Upper
130 100
95 97.1 0.0 202.6





#45

1,2-Dichloropropane

Concen: 49.405 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 37161

Ion Ratio Lower Upper

63 100

65 30.9 23.8 35.6



#46

Dibromomethane

Concen: 49.917 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39




Tgt Ion: 93 Resp: 26968


Ion Ratio Lower Upper

93 100

95 84.2 66.1 99.1

174 109.3 78.7 118.1





#47

Bromodichloromethane

Concen: 50.307 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. -0.000 min

Lab File: VX027263.D





Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:	83	Resp:	53048
Ion	Ratio	Lower	Upper
83	100		
85	64.4	51.4	77.0
127	8.7	6.9	10.3



#48

Methyl methacrylate

Concen: 49.539 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. -0.000 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Tgt Ion:	41	Resp:	42635
Ion	Ratio	Lower	Upper
41	100		
69	87.0	62.2	93.4
39	56.3	45.4	68.0





#49

1,4-Dioxane

Concen: 991.189 ug/l

RT: 7.659 min Scan# 1079

Delta R.T. -0.006 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 19285

Ion Ratio Lower Upper

88 100

43 32.1 29.9 44.9

58 68.4 57.8 86.6



#50

Toluene-d8

Concen: 48.915 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX027263.D




Acq: 02 Mar 2022 12:39

Tgt Ion: 98 Resp: 131925

Ion Ratio Lower Upper

98 100

100 66.4 53.2 79.8





#51

4-Methyl-2-Pentanone

Concen: 254.260 ug/l

RT: 8.573 min Scan# 11

Delta R.T. -0.007 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 293834

Ion Ratio Lower Upper

43 100

58 39.5 30.5 45.7



#52

Toluene

Concen: 51.524 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX027263.D




Acq: 02 Mar 2022 12:39


Tgt Ion: 92 Resp: 101565

Ion Ratio Lower Upper

92 100

91 167.2 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 46.107 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 54722

Ion Ratio Lower Upper

75 100

77 32.1 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 51.769 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39




Tgt Ion: 75 Resp: 60936

Ion Ratio Lower Upper

75 100

77 32.1 26.2 39.2

39 48.4 43.5 65.3









#55
1,1,2-Trichloroethane
Concen: 51.650 ug/l
RT: 9.153 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 97 Resp: 40326




Ion	Ratio	Lower	Upper
97	100		
83	82.7	68.4	102.6
85	53.1	44.0	66.0
99	61.0	49.8	74.6




#56
Ethyl methacrylate
Concen: 51.016 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion: 69 Resp: 62013

Ion	Ratio	Lower	Upper
69	100		
41	64.2	57.2	85.8
39	36.0	30.9	46.3









#57
1,3-Dichloropropane
Concen: 50.793 ug/l
RT: 9.311 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 76 Resp: 65961
Ion Ratio Lower Upper
76 100
78 32.4 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 258.226 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion: 63 Resp: 151300
Ion Ratio Lower Upper
63 100
106 31.5 23.0 34.6





#59

2-Hexanone

Concen: 248.487 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 219478

Ion Ratio Lower Upper

43 100

58 55.6 26.4 79.2



#60

Dibromochloromethane

Concen: 53.287 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX027263.D




Acq: 02 Mar 2022 12:39


Tgt Ion: 129 Resp: 43159

Ion Ratio Lower Upper

129 100

127 75.7 38.5 115.5









#61
1,2-Dibromoethane
Concen: 51.138 ug/l
RT: 9.610 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 42520
Ion Ratio Lower Upper
107 100
109 93.8 75.5 113.3



#62
4-Bromofluorobenzene
Concen: 50.485 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39


Tgt Ion: 95 Resp: 52265
Ion Ratio Lower Upper
95 100
174 79.2 0.0 148.6
176 79.3 0.0 141.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:117 Resp: 107169
Ion Ratio Lower Upper
117 100
82 54.6 45.2 67.8
119 32.6 26.2 39.2



#64
Tetrachloroethene
Concen: 45.793 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39


Tgt Ion:164 Resp: 44401
Ion Ratio Lower Upper
164 100
166 127.1 101.8 152.6
129 100.9 83.3 124.9
131 88.5 78.0 117.0








#65
Chlorobenzene
Concen: 51.034 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:112 Resp: 111489
Ion Ratio Lower Upper
112 100
114 32.8 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 51.778 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion:131 Resp: 41020
Ion Ratio Lower Upper
131 100
133 94.3 47.9 143.8
119 65.2 31.7 95.1





#67

Ethyl Benzene

Concen: 51.543 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 197668

Ion Ratio Lower Upper

91 100

106 32.0 24.7 37.1



#68

m/p-Xylenes

Concen: 101.090 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. -0.000 min

Lab File: VX027263.D




Acq: 02 Mar 2022 12:39


Tgt Ion:106 Resp: 149041

Ion Ratio Lower Upper

106 100

91 202.3 165.3 247.9









#69
o-Xylene
Concen: 50.189 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:106 Resp: 74265
Ion Ratio Lower Upper
106 100
91 211.7 109.7 329.0



#70
Styrene
Concen: 52.014 ug/l
RT: 10.658 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion:104 Resp: 125940
Ion Ratio Lower Upper
104 100
78 49.3 42.5 63.7
103 54.3 43.3 64.9









#71
Bromoform
Concen: 46.844 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:173 Resp: 31867
Ion Ratio Lower Upper
173 100
175 47.1 24.3 72.9
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion:152 Resp: 53157
Ion Ratio Lower Upper
152 100
115 84.8 44.2 132.6
150 178.2 0.0 348.2









#73
Isopropylbenzene
Concen: 51.045 ug/l
RT: 10.963 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:105 Resp: 194160
Ion Ratio Lower Upper
105 100
120 26.9 13.1 39.1



#74
N-ethyl acetate
Concen: 52.049 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion: 43 Resp: 75328
Ion Ratio Lower Upper
43 100
70 44.9 32.9 49.3
55 25.4 18.3 27.5
61 25.2 18.9 28.3









#75
 1,1,2,2-Tetrachloroethane
 Concen: 50.865 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX027263.D
 Acq: 02 Mar 2022 12:39




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 83 Resp: 57146
 Ion Ratio Lower Upper
 83 100
 131 11.2 5.3 15.9
 85 65.3 32.0 96.2



#76
 1,2,3-Trichloropropane
 Concen: 62.096 ug/l
 RT: 11.244 min Scan# 1666
 Delta R.T. -0.000 min
 Lab File: VX027263.D
 Acq: 02 Mar 2022 12:39


Tgt Ion: 75 Resp: 66487
 Ion Ratio Lower Upper
 75 100
 77 30.8 18.3 54.8








#77
Bromobenzene
Concen: 50.447 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:156 Resp: 47862
Ion Ratio Lower Upper
156 100
77 155.8 84.5 253.4
158 97.0 48.1 144.4



#78
n-propylbenzene
Concen: 50.598 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion: 91 Resp: 218745
Ion Ratio Lower Upper
91 100
120 23.9 11.6 34.7









#79
2-Chlorotoluene
Concen: 50.830 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 91 Resp: 134659
Ion Ratio Lower Upper
91 100
126 34.8 16.4 49.2



#80
1,3,5-Trimethylbenzene
Concen: 51.407 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion: 105 Resp: 164420
Ion Ratio Lower Upper
105 100
120 49.1 24.0 72.0









#81
trans-1,4-Dichloro-2-butene
Concen: 45.220 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 16360
Ion Ratio Lower Upper
75 100
53 104.2 94.0 141.0
89 47.4 37.0 55.4



#82
4-Chlorotoluene
Concen: 50.647 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion: 91 Resp: 153278
Ion Ratio Lower Upper
91 100
126 30.2 14.1 42.4









#83
tert-Butylbenzene
Concen: 51.909 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.006 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 161439
Ion Ratio Lower Upper
119 100
91 58.4 31.4 94.0
134 23.8 11.8 35.3



#84
1,2,4-Trimethylbenzene
Concen: 52.505 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion:105 Resp: 167041
Ion Ratio Lower Upper
105 100
120 45.3 21.9 65.5









#85
 sec-Butylbenzene
 Concen: 53.786 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX027263.D
 Acq: 02 Mar 2022 12:39




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 209744
 Ion Ratio Lower Upper
 105 100
 134 20.9 10.1 30.1



#86
 p-Isopropyltoluene
 Concen: 54.195 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX027263.D
 Acq: 02 Mar 2022 12:39

Tgt Ion:119 Resp: 180168
 Ion Ratio Lower Upper
 119 100
 134 26.3 13.0 39.0
 91 22.8 12.6 37.6





#87

1,3-Dichlorobenzene

Concen: 51.184 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 91377

Ion Ratio Lower Upper

146 100

111 40.2 21.3 63.7

148 64.6 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 49.711 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39




Tgt Ion:146 Resp: 90741


Ion Ratio Lower Upper

146 100

111 39.2 20.2 60.6

148 63.7 31.4 94.0





#89

n-Butylbenzene

Concen: 52.253 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 145191

Ion Ratio Lower Upper

91 100

92 54.9 27.2 81.6

134 27.4 13.1 39.1



#90

Hexachloroethane

Concen: 52.814 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX027263.D




Acq: 02 Mar 2022 12:39


Tgt Ion: 117 Resp: 27197

Ion Ratio Lower Upper

117 100

201 76.6 35.6 106.8





#91

1,2-Dichlorobenzene

Concen: 50.666 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 88256

Ion Ratio Lower Upper

146 100

111 41.8 21.8 65.3

148 63.6 32.0 96.2



#92

1,2-Dibromo-3-Chloropropane

Concen: 51.608 ug/l

RT: 12.945 min Scan# 1945

Delta R.T. -0.000 min

Lab File: VX027263.D

Acq: 02 Mar 2022 12:39




Tgt Ion: 75 Resp: 13461

Ion Ratio Lower Upper

75 100

155 92.3 40.0 119.9


157 116.0 51.6 154.8








#93
1,2,4-Trichlorobenzene
Concen: 51.719 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:180 Resp: 56850
Ion Ratio Lower Upper
180 100
182 95.9 46.7 140.0
145 31.7 16.3 48.8




#94
Hexachlorobutadiene
Concen: 51.460 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39




Tgt Ion:225 Resp: 23484
Ion Ratio Lower Upper
225 100
223 61.7 32.0 96.0
227 63.9 32.3 96.9








#95
Naphthalene
Concen: 52.219 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 195049
Ion Ratio Lower Upper
128 100
127 12.9 10.3 15.5
129 10.8 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 52.799 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX027263.D
Acq: 02 Mar 2022 12:39

Tgt Ion:180 Resp: 58174
Ion Ratio Lower Upper
180 100
182 96.1 47.2 141.6
145 34.2 17.5 52.5

