

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030222\
 Data File : VX027265.D
 Acq On : 02 Mar 2022 13:35
 Operator : JC/MD
 Sample : VX0302WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0302WBL01

Quant Time: Mar 03 00:05:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	82124	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	138071	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58752	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51146	51.856	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.720%	
35) Dibromofluoromethane	5.391	113	43918	51.244	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.480%	
50) Toluene-d8	8.653	98	159251	49.695	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.400%	
62) 4-Bromofluorobenzene	11.079	95	59338	48.240	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.480%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030222\
Data File : VX027265.D
Acq On : 02 Mar 2022 13:35
Operator : JC/MD
Sample : VX0302WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0302WBL01

Quant Time: Mar 03 00:05:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

