

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030222\
 Data File : VX027268.D
 Acq On : 02 Mar 2022 17:27
 Operator : JC/MD
 Sample : PB143060TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB143060TB

Quant Time: Mar 03 00:07:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82394	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	136227	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	49730	50.255	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.500%
35) Dibromofluoromethane	5.385	113	43363	51.282	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.560%
50) Toluene-d8	8.647	98	157373	49.774	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.540%
62) 4-Bromofluorobenzene	11.079	95	59011	48.623	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.240%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	7249	16.685	ug/l	# 88
20) Methylene Chloride	2.788	84	33980	37.734	ug/l	92
43) Isopropyl Acetate	6.342	43	7099	3.406	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030222\
Data File : VX027268.D
Acq On : 02 Mar 2022 17:27
Operator : JC/MD
Sample : PB143060TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB143060TB

Quant Time: Mar 03 00:07:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

