

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030222\
 Data File : VX027270.D
 Acq On : 02 Mar 2022 18:14
 Operator : JC/MD
 Sample : N1757-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 34064

Quant Time: Mar 03 00:07:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	80807	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	132443	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120976	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61459	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47590	49.037	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.080%
35) Dibromofluoromethane	5.391	113	41986	51.072	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.140%
50) Toluene-d8	8.653	98	153904	50.068	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.140%
62) 4-Bromofluorobenzene	11.085	95	58688	49.739	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.480%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	183143	429.807	ug/l	95
64) Tetrachloroethene	9.275	164	16910	15.450	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030222\
Data File : VX027270.D
Acq On : 02 Mar 2022 18:14
Operator : JC/MD
Sample : N1757-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
34064

Quant Time: Mar 03 00:07:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

