

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030222\
 Data File : VX027271.D
 Acq On : 02 Mar 2022 18:37
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 03 00:07:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.497	0.420	15.5	74	0.00
3 P	Chloromethane	0.465	0.362	22.2	67	0.00
4 C	Vinyl Chloride	0.463	0.363	21.6#	66	0.00
5 T	Bromomethane	0.210	0.161	23.3	75	0.00
6 T	Chloroethane	0.284	0.235	17.3	72	0.00
7 T	Trichlorofluoromethane	0.720	0.666	7.5	79	0.00
8 T	Diethyl Ether	0.251	0.216	13.9	72	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.501	0.463	7.6	79	0.00
10 T	Methyl Iodide	0.673	0.493	26.7#	60	0.00
11 T	Tert butyl alcohol	0.117	0.104	11.1	77	0.00
12 CM	1,1-Dichloroethene	0.487	0.444	8.8#	78	0.00
13 T	Acrolein	0.108	0.104	3.7	86	0.00
14 T	Allyl chloride	0.857	0.765	10.7	75	0.00
15 T	Acrylonitrile	0.300	0.287	4.3	80	0.00
16 T	Acetone	0.264	0.235	11.0	79	0.00
17 T	Carbon Disulfide	1.266	1.067	15.7	72	0.00
18 T	Methyl Acetate	0.681	0.622	8.7	77	0.00
19 T	Methyl tert-butyl Ether	1.670	1.607	3.8	80	0.00
20 T	Methylene Chloride	0.546	0.516	5.5	82	0.00
21 T	trans-1,2-Dichloroethene	0.519	0.481	7.3	78	0.00
22 T	Diisopropyl ether	1.671	1.559	6.7	78	0.00
23 T	Vinyl Acetate	1.411	1.366	3.2	79	0.00
24 P	1,1-Dichloroethane	0.944	0.886	6.1	80	0.00
25 T	2-Butanone	0.420	0.388	7.6	78	0.00
26 T	2,2-Dichloropropane	0.685	0.676	1.3	82	0.00
27 T	cis-1,2-Dichloroethene	0.602	0.580	3.7	81	0.00
28 T	Bromochloromethane	0.382	0.374	2.1	81	0.00
29 T	Tetrahydrofuran	0.281	0.255	9.3	77	0.00
30 C	Chloroform	0.981	0.932	5.0#	81	0.00
31 T	Cyclohexane	0.899	0.787	12.5	75	0.00
32 T	1,1,1-Trichloroethane	0.849	0.829	2.4	81	0.00
33 S	1,2-Dichloroethane-d4	0.601	0.565	6.0	84	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.310	0.300	3.2	84	0.00
36 T	1,1-Dichloropropene	0.453	0.427	5.7	78	0.00
37 T	Ethyl Acetate	0.497	0.465	6.4	79	0.00
38 T	Carbon Tetrachloride	0.472	0.467	1.1	81	0.00
39 T	Methylcyclohexane	0.563	0.540	4.1	80	0.00
40 TM	Benzene	1.325	1.266	4.5	79	0.00
41 T	Methacrylonitrile	0.273	0.261	4.4	79	0.00
42 TM	1,2-Dichloroethane	0.470	0.462	1.7	82	0.00
43 T	Isopropyl Acetate	0.765	0.753	1.6	80	0.00
44 TM	Trichloroethene	0.396	0.378	4.5	79	0.00
45 C	1,2-Dichloropropane	0.324	0.330	-1.9#	83	0.00
46 T	Dibromomethane	0.232	0.232	0.0	81	0.00
47 T	Bromodichloromethane	0.454	0.451	0.7	82	0.00
48 T	Methyl methacrylate	0.370	0.376	-1.6	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030222\
 Data File : VX027271.D
 Acq On : 02 Mar 2022 18:37
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 03 00:07:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	80	0.00
50 S	Toluene-d8	1.160	1.099	5.3	82	0.00
51 T	4-Methyl-2-Pentanone	0.497	0.496	0.2	81	0.00
52 CM	Toluene	0.848	0.829	2.2#	81	0.00
53 T	t-1,3-Dichloropropene	0.446	0.500	-12.1	87	0.00
54 T	cis-1,3-Dichloropropene	0.506	0.529	-4.5	83	0.00
55 T	1,1,2-Trichloroethane	0.336	0.344	-2.4	84	0.00
56 T	Ethyl methacrylate	0.523	0.537	-2.7	81	0.00
57 T	1,3-Dichloropropane	0.559	0.578	-3.4	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.252	0.261	-3.6	83	0.00
59 T	2-Hexanone	0.380	0.392	-3.2	83	0.00
60 T	Dibromochloromethane	0.348	0.390	-12.1	89	0.00
61 T	1,2-Dibromoethane	0.358	0.366	-2.2	82	0.00
62 S	4-Bromofluorobenzene	0.445	0.428	3.8	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.452	0.418	7.5	78	0.00
65 PM	Chlorobenzene	1.019	0.986	3.2	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.382	-3.2	86	0.00
67 C	Ethyl Benzene	1.789	1.725	3.6#	81	0.00
68 T	m/p-Xylenes	0.688	0.675	1.9	81	0.00
69 T	o-Xylene	0.690	0.667	3.3	81	0.00
70 T	Styrene	1.130	1.132	-0.2	81	0.00
71 P	Bromoform	0.280	0.295	-5.4	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.578	3.493	2.4	82	0.00
74 T	N-amyl acetate	1.361	1.423	-4.6	83	0.00
75 P	1,1,2,2-Tetrachloroethane	1.057	1.064	-0.7	87	0.00
76 T	1,2,3-Trichloropropane	1.007	1.025	-1.8	85	0.00
77 T	Bromobenzene	0.892	0.883	1.0	83	0.00
78 T	n-propylbenzene	4.066	4.028	0.9	82	0.00
79 T	2-Chlorotoluene	2.492	2.384	4.3	81	0.00
80 T	1,3,5-Trimethylbenzene	3.008	3.010	-0.1	82	0.00
81 T	trans-1,4-Dichloro-2-butene	0.315	0.302	4.1	80	0.00
82 T	4-Chlorotoluene	2.847	2.903	-2.0	85	0.00
83 T	tert-Butylbenzene	2.925	2.897	1.0	83	0.00
84 T	1,2,4-Trimethylbenzene	2.993	2.947	1.5	81	0.00
85 T	sec-Butylbenzene	3.668	3.658	0.3	82	0.00
86 T	p-Isopropyltoluene	3.127	3.075	1.7	81	0.00
87 T	1,3-Dichlorobenzene	1.679	1.659	1.2	83	0.00
88 T	1,4-Dichlorobenzene	1.717	1.687	1.7	84	0.00
89 T	n-Butylbenzene	2.614	2.713	-3.8	84	0.00
90 T	Hexachloroethane	0.484	0.508	-5.0	84	0.00
91 T	1,2-Dichlorobenzene	1.638	1.675	-2.3	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.245	0.257	-4.9	85	0.00
93 T	1,2,4-Trichlorobenzene	1.034	1.067	-3.2	85	0.00
94 T	Hexachlorobutadiene	0.429	0.443	-3.3	86	0.00
95 T	Naphthalene	3.513	3.654	-4.0	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030222\
Data File : VX027271.D
Acq On : 02 Mar 2022 18:37
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 03 00:07:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.036	1.121	-8.2	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6