

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030223\
 Data File : VX034297.D
 Acq On : 02 Mar 2023 10:13
 Operator : JC/MD
 Sample : VX0302MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0302MBL01

Quant Time: Mar 03 00:19:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	292664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	484191	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	427428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	190423	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	174607	44.300	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.600%
35) Dibromofluoromethane	5.379	113	153338	48.639	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.280%
50) Toluene-d8	8.646	98	584426	51.633	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.260%
62) 4-Bromofluorobenzene	11.079	95	199126	45.993	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030223\
Data File : VX034297.D
Acq On : 02 Mar 2023 10:13
Operator : JC/MD
Sample : VX0302MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0302MBL01

Quant Time: Mar 03 00:19:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

