

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030223\
 Data File : VX034303.D
 Acq On : 02 Mar 2023 12:46
 Operator : JC/MD
 Sample : 01759-01 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DEP-1

Quant Time: Mar 03 00:22:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	256315	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	416598	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	357430	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146786	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149748	43.381	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.760%
35) Dibromofluoromethane	5.385	113	129433	47.718	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.440%
50) Toluene-d8	8.647	98	493895	50.715	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.420%
62) 4-Bromofluorobenzene	11.079	95	159193	42.735	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030223\
Data File : VX034303.D
Acq On : 02 Mar 2023 12:46
Operator : JC/MD
Sample : 01759-01 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DEP-1

Quant Time: Mar 03 00:22:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

