

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030223\
 Data File : VX034305.D
 Acq On : 02 Mar 2023 14:04
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 03 00:23:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	265031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	438961	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	394312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	180869	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	159558	44.702	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.400%
35) Dibromofluoromethane	5.385	113	137325	48.048	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.100%
50) Toluene-d8	8.647	98	529846	51.635	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.260%
62) 4-Bromofluorobenzene	11.079	95	183183	46.670	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.340%
Target Compounds						
20) Methylene Chloride	2.794	84	1721	0.519	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030223\
Data File : VX034305.D
Acq On : 02 Mar 2023 14:04
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 03 00:23:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

