

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030322\
 Data File : VX027284.D
 Acq On : 03 Mar 2022 21:23
 Operator : JC/MD
 Sample : N1769-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20220302

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/04/2022
 Supervised By : Mahesh Dadoda 03/04/2022

Quant Time: Mar 04 02:42:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030322W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 04 01:00:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	11922	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	64642	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	57449	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	25359	30.725	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.400%
60) 4-Bromofluorobenzene	11.085	95	25663	27.574	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.900%
63) Toluene-d8	8.653	98	74379	29.616	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.733%
Target Compounds						
15) Acetone	2.386	58	661m	5.748	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030322\
Data File : VX027284.D
Acq On : 03 Mar 2022 21:23
Operator : JC/MD
Sample : N1769-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20220302

Quant Time: Mar 04 02:42:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030322W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 04 01:00:22 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 03/04/2022
Supervised By : Mahesh Dadoda 03/04/2022

