

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030322\
 Data File : VX027287.D
 Acq On : 03 Mar 2022 22:34
 Operator : JC/MD
 Sample : N1728-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 002-35TH-AVE(MAR)

Quant Time: Mar 04 02:42:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030322W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 04 01:00:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	12244	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	66044	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	59631	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	25885	30.537	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 101.800%			
60) 4-Bromofluorobenzene	11.079	95	27046	27.997	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 93.333%			
63) Toluene-d8	8.653	98	77422	29.699	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 99.000%			
Target Compounds						
					Qvalue	
15) Acetone	2.380	58	7861	66.560	ug/l	99
16) Carbon Disulfide	2.526	76	3111	1.671	ug/l	96
30) 2-Butanone	4.568	43	3620	6.352	ug/l #	93
58) 4-Methyl-2-Pentanone	8.574	43	3103	2.877	ug/l	97
62) Toluene	8.726	91	2752	0.847	ug/l	95
67) m/p-Xylenes	10.305	106	347	0.242	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030322\
Data File : VX027287.D
Acq On : 03 Mar 2022 22:34
Operator : JC/MD
Sample : N1728-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
002-35TH-AVE(MAR)

Quant Time: Mar 04 02:42:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030322W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 04 01:00:22 2022
Response via : Initial Calibration

