


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030322\
 Data File : VX027272.D
 Acq On : 03 Mar 2022 14:10
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030322W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Fri Mar 04 01:00:22 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	3573	PASS
75	95	30	60	49.2	9911	PASS
95	95	100	100	100.0	20140	PASS
96	95	5	9	6.9	1380	PASS
173	174	0.00	2	1.2	188	PASS
174	95	50	100	76.1	15324	PASS
175	174	5	9	7.6	1168	PASS
176	174	95	101	99.0	15172	PASS
177	176	5	9	6.9	1052	PASS