

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030323\
 Data File : VX034320.D
 Acq On : 03 Mar 2023 14:13
 Operator : JC/MD
 Sample : PB151167ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151167ZHE#01

Quant Time: Mar 03 22:12:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	296840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	503772	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	459938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	205948	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	190495	47.651	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	95.300%	
35) Dibromofluoromethane	5.385	113	161920	49.365	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.740%	
50) Toluene-d8	8.647	98	613935	52.132	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	104.260%	
62) 4-Bromofluorobenzene	11.079	95	222984	49.502	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	99.000%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.941	59	6653	7.441	ug/l #	96
16) Acetone	2.380	43	69983	33.936	ug/l	96
20) Methylene Chloride	2.794	84	13098	3.525	ug/l	97
37) Ethyl Acetate	4.715	43	90036	16.608	ug/l	99
43) Isopropyl Acetate	6.336	43	360424	37.675	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030323\
Data File : VX034320.D
Acq On : 03 Mar 2023 14:13
Operator : JC/MD
Sample : PB151167ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151167ZHE#01

Quant Time: Mar 03 22:12:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

