

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030323\
 Data File : VX034321.D
 Acq On : 03 Mar 2023 14:37
 Operator : JC/MD
 Sample : PB151167ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151167ZHE#03

Quant Time: Mar 03 22:12:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	284694	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	483159	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	435396	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	192329	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	176666	46.077	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.160%		
35) Dibromofluoromethane	5.385	113	152179	48.375	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.740%		
50) Toluene-d8	8.647	98	583222	51.637	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	103.280%		
62) 4-Bromofluorobenzene	11.079	95	208058	48.159	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.320%		
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.941	59	6736	7.855	ug/l #	92
16) Acetone	2.380	43	62040	31.367	ug/l	95
20) Methylene Chloride	2.794	84	11198	3.142	ug/l	95
37) Ethyl Acetate	4.715	43	81843	15.741	ug/l	100
43) Isopropyl Acetate	6.336	43	334446	36.451	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030323\
Data File : VX034321.D
Acq On : 03 Mar 2023 14:37
Operator : JC/MD
Sample : PB151167ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151167ZHE#03

Quant Time: Mar 03 22:12:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

