

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030323\
 Data File : VX034322.D
 Acq On : 03 Mar 2023 15:00
 Operator : JC/MD
 Sample : PB151167ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151167ZHE#04

Quant Time: Mar 03 22:13:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	272873	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	455017	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	418994	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	179536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	168102	45.743	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.480%
35) Dibromofluoromethane	5.385	113	144507	48.777	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.560%
50) Toluene-d8	8.647	98	558045	52.464	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.920%
62) 4-Bromofluorobenzene	11.079	95	192560	47.328	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.660%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	6037	7.345	ug/l #	94
16) Acetone	2.380	43	61327	32.350	ug/l	92
20) Methylene Chloride	2.788	84	11255	3.295	ug/l	95
37) Ethyl Acetate	4.715	43	79588	16.254	ug/l	99
43) Isopropyl Acetate	6.336	43	329940	38.184	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030323\
Data File : VX034322.D
Acq On : 03 Mar 2023 15:00
Operator : JC/MD
Sample : PB151167ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151167ZHE#04

Quant Time: Mar 03 22:13:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

