

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030323\
 Data File : VX034324.D
 Acq On : 03 Mar 2023 15:47
 Operator : JC/MD
 Sample : PB151167ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151167ZHE#02

Quant Time: Mar 03 22:14:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	275307	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	462452	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	431438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	192502	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	170229	45.912	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	91.820%		
35) Dibromofluoromethane	5.385	113	144936	48.135	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.280%		
50) Toluene-d8	8.647	98	566921	52.441	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	104.880%		
62) 4-Bromofluorobenzene	11.079	95	199269	48.190	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.380%		
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.940	59	6519	7.861	ug/l #	91
16) Acetone	2.380	43	61129	31.961	ug/l	100
20) Methylene Chloride	2.788	84	11644	3.379	ug/l	97
37) Ethyl Acetate	4.714	43	79086	15.891	ug/l	99
43) Isopropyl Acetate	6.336	43	323383	36.823	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030323\
Data File : VX034324.D
Acq On : 03 Mar 2023 15:47
Operator : JC/MD
Sample : PB151167ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151167ZHE#02

Quant Time: Mar 03 22:14:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

