

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030323\
 Data File : VX034330.D
 Acq On : 03 Mar 2023 18:09
 Operator : JC/MD
 Sample : 01767-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CE-54-030123

Quant Time: Mar 03 22:16:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	260679	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	441522	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	401903	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	173880	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	159673	45.481	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.960%
35) Dibromofluoromethane	5.385	113	138364	48.131	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.260%
50) Toluene-d8	8.647	98	535308	51.864	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.720%
62) 4-Bromofluorobenzene	11.079	95	186905	47.342	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.680%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	55203	30.482	ug/l	90
20) Methylene Chloride	2.788	84	9970	3.055	ug/l	96
37) Ethyl Acetate	4.715	43	72269	15.210	ug/l	99
43) Isopropyl Acetate	6.342	43	300214	35.805	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030323\
Data File : VX034330.D
Acq On : 03 Mar 2023 18:09
Operator : JC/MD
Sample : 01767-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CE-54-030123

Quant Time: Mar 03 22:16:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

