

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
 Data File : VX045105.D
 Acq On : 03 Mar 2025 11:30
 Operator : JC/MD
 Sample : PB166928TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB166928TB

Quant Time: Mar 04 05:20:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	75337	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152031	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	143120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	62542	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	62450	52.114	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.220%	
35) Dibromofluoromethane	5.379	113	51016	50.183	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.360%	
50) Toluene-d8	8.641	98	189518	51.423	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.840%	
62) 4-Bromofluorobenzene	11.079	95	66977	54.842	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 109.680%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	23995	43.157	ug/l	97
18) Methyl Acetate	2.703	43	1459	1.091	ug/l	95
20) Methylene Chloride	2.782	84	2123	1.928	ug/l #	92
43) Isopropyl Acetate	6.336	43	62269	23.441	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
Data File : VX045105.D
Acq On : 03 Mar 2025 11:30
Operator : JC/MD
Sample : PB166928TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB166928TB

Quant Time: Mar 04 05:20:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

