

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
 Data File : VX045106.D
 Acq On : 03 Mar 2025 11:53
 Operator : JC/MD
 Sample : PB166928ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB166928ZHE#01

Quant Time: Mar 04 05:21:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	81781	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	164951	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	150327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	64106	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	67300	51.736	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.480%		
35) Dibromofluoromethane	5.373	113	55525	50.341	ug/l	-0.01
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.680%		
50) Toluene-d8	8.641	98	206076	51.536	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.080%		
62) 4-Bromofluorobenzene	11.079	95	70419	53.144	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 106.280%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	24490	40.576	ug/l	99
18) Methyl Acetate	2.697	43	1867	1.286	ug/l	91
20) Methylene Chloride	2.782	84	1769	1.480	ug/l	95
43) Isopropyl Acetate	6.336	43	61044	21.180	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
Data File : VX045106.D
Acq On : 03 Mar 2025 11:53
Operator : JC/MD
Sample : PB166928ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB166928ZHE#01

Quant Time: Mar 04 05:21:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

