

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
 Data File : VX045108.D
 Acq On : 03 Mar 2025 13:23
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 04 05:22:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	81581	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.750	114	165592	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	153477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68603	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	66756	51.443	ug/l	-0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.880%
35) Dibromofluoromethane	5.373	113	54967	49.642	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.280%
50) Toluene-d8	8.640	98	203220	50.625	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.240%
62) 4-Bromofluorobenzene	11.079	95	73061	54.925	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	109.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
Data File : VX045108.D
Acq On : 03 Mar 2025 13:23
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 04 05:22:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

