

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
 Data File : VX045109.D
 Acq On : 03 Mar 2025 13:46
 Operator : JC/MD
 Sample : Q1478-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-AQ-DRUM-610-022825

Quant Time: Mar 04 05:22:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	84810	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	169281	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138779	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	67488	50.027	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.060%	
35) Dibromofluoromethane	5.379	113	55945	49.424	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.840%	
50) Toluene-d8	8.641	98	176471	43.003	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 86.000%	
62) 4-Bromofluorobenzene	11.079	95	69546	51.143	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 102.280%	
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	3854	2.974	ug/l	89
9) 1,1,2-Trichlorotrifluo...	2.319	101	1342729	1362.203	ug/l	99
12) 1,1-Dichloroethene	2.313	96	28456	27.305	ug/l	98
16) Acetone	2.380	43	3967	6.338	ug/l	96
18) Methyl Acetate	2.709	43	806	0.535	ug/l #	84
21) trans-1,2-Dichloroethene	3.087	96	53998	52.448	ug/l	93
24) 1,1-Dichloroethane	3.599	63	24136	11.415	ug/l	98
27) cis-1,2-Dichloroethene	4.477	96	1844339	1451.156	ug/l	96
30) Chloroform	5.080	83	6975	3.320	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	53454	31.503	ug/l	98
40) Benzene	6.031	78	5946	1.213	ug/l	100
42) 1,2-Dichloroethane	6.086	62	4777	2.706	ug/l	95
44) Trichloroethene	7.159	130	14962403	12999.289	ug/l	95
52) Toluene	8.714	92	9358	3.254	ug/l	100
55) 1,1,2-Trichloroethane	9.153	97	504	0.426	ug/l #	78
64) Tetrachloroethene	9.269	164	732498	824.121	ug/l	98
68) m/p-Xylenes	10.299	106	651	0.348	ug/l	86
88) 1,4-Dichlorobenzene	12.036	146	15569	6.838	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
Data File : VX045109.D
Acq On : 03 Mar 2025 13:46
Operator : JC/MD
Sample : Q1478-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-AQ-DRUM-610-022825

Quant Time: Mar 04 05:22:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

