

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030419\
 Data File : VX007813.D
 Acq On : 04 Mar 2019 16:22
 Operator : JC/SP
 Sample : K1719-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-GW-485-487

Quant Time: Mar 04 22:15:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	337013	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	523490	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	489380	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	236826	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	183687	44.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.00%	
35) Dibromofluoromethane	5.51	113	166667	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
50) Toluene-d8	8.71	98	629735	47.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.02%	
62) 4-Bromofluorobenzene	11.13	95	232201	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
Target Compounds						
16) Acetone	2.45	43	12235	6.591	ug/l	97
25) 2-Butanone	4.70	43	4882	1.726	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	1522	0.373	ug/l	97
44) Trichloroethene	7.21	130	29223	7.312	ug/l	89
64) Tetrachloroethene	9.34	164	4467	1.124	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030419\
Data File : VX007813.D
Acq On : 04 Mar 2019 16:22
Operator : JC/SP
Sample : K1719-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-GW-485-487

Quant Time: Mar 04 22:15:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

