

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030419\
 Data File : VX007824.D
 Acq On : 04 Mar 2019 21:15
 Operator : JC/SP
 Sample : K1707-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S11

Quant Time: Mar 04 22:52:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	337579	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	522716	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	498544	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	237874	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	185627	44.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.78%	
35) Dibromofluoromethane	5.51	113	172592	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
50) Toluene-d8	8.71	98	641412	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
62) 4-Bromofluorobenzene	11.13	95	231789	47.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.42%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	11610	6.244	ug/l	98
18) Methyl Acetate	2.78	43	14967	3.689	ug/l	96
20) Methylene Chloride	2.86	84	2091639	609.157	ug/l	94
25) 2-Butanone	4.71	43	3125	1.103	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	4975	1.218	ug/l	97
43) Isopropyl Acetate	6.46	43	17341	2.158	ug/l	99
64) Tetrachloroethene	9.34	164	19020	4.699	ug/l	97
95) Naphthalene	13.83	128	18789	1.200	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030419\
Data File : VX007824.D
Acq On : 04 Mar 2019 21:15
Operator : JC/SP
Sample : K1707-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S11

Quant Time: Mar 04 22:52:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

