

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030421\
 Data File : VX020968.D
 Acq On : 02 Mar 2021 16:48
 Operator : JC/MD
 Sample : VX0304WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0304WBL01

Quant Time: Mar 03 01:15:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	205959	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	369519	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	360864	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	152371	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	131452	47.25	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.50%
35) Dibromofluoromethane	5.464	113	115430	51.41	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.82%
50) Toluene-d8	8.694	98	455296	52.01	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.02%
62) 4-Bromofluorobenzene	11.118	95	165407	51.10	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030421\
Data File : VX020968.D
Acq On : 02 Mar 2021 16:48
Operator : JC/MD
Sample : VX0304WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0304WBL01

Quant Time: Mar 03 01:15:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

