

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030421\
 Data File : VX020974.D
 Acq On : 02 Mar 2021 19:05
 Operator : JC/MD
 Sample : M1535-01 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW807B-15-20-022721

Quant Time: Mar 03 01:17:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	201667	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	370099	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	361438	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	159530	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	129978	47.72	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.44%
35) Dibromofluoromethane	5.464	113	115387	51.31	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.62%
50) Toluene-d8	8.694	98	456763	52.10	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.20%
62) 4-Bromofluorobenzene	11.118	95	170012	52.44	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.88%
Target Compounds						
16) Acetone	2.417	43	3749	3.34	ug/l	Qvalue 91
20) Methylene Chloride	2.840	84	1360	0.51	ug/l #	86
30) Chloroform	5.182	83	9562	2.17	ug/l	96
47) Bromodichloromethane	7.870	83	1106	0.30	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030421\
Data File : VX020974.D
Acq On : 02 Mar 2021 19:05
Operator : JC/MD
Sample : M1535-01 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW807B-15-20-022721

Quant Time: Mar 03 01:17:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

