

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030421\
 Data File : VX020975.D
 Acq On : 02 Mar 2021 19:28
 Operator : JC/MD
 Sample : M1545-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-725

Quant Time: Mar 03 01:18:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	195548	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	355318	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	346113	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	155965	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	126249	47.80	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.60%
35) Dibromofluoromethane	5.464	113	110946	51.39	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.78%
50) Toluene-d8	8.694	98	435359	51.72	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.44%
62) 4-Bromofluorobenzene	11.118	95	168019	53.98	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.96%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.999	59	987	1.68	ug/l	# 93
16) Acetone	2.417	43	2154	1.98	ug/l	99
20) Methylene Chloride	2.834	84	661	0.25	ug/l	94
68) m/p-Xylenes	10.341	106	2219	0.46	ug/l	89
69) o-Xylene	10.683	106	1618	0.34	ug/l	87
95) Naphthalene	13.818	128	9917	0.87	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030421\
Data File : VX020975.D
Acq On : 02 Mar 2021 19:28
Operator : JC/MD
Sample : M1545-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-725

Quant Time: Mar 03 01:18:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

