

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030421\
 Data File : VX020977.D
 Acq On : 02 Mar 2021 20:14
 Operator : JC/MD
 Sample : M1545-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-F

Quant Time: Mar 03 01:18:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	200345	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	365724	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	360066	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	165923	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	130160	48.10	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.20%
35) Dibromofluoromethane	5.464	113	113505	51.08	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.16%
50) Toluene-d8	8.694	98	455151	52.53	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.06%
62) 4-Bromofluorobenzene	11.118	95	177368	55.36	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.72%
Target Compounds						
16) Acetone	2.417	43	5209	4.67	ug/l	95
20) Methylene Chloride	2.840	84	1576	0.59	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030421\
Data File : VX020977.D
Acq On : 02 Mar 2021 20:14
Operator : JC/MD
Sample : M1545-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-F

Quant Time: Mar 03 01:18:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

