

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030421\
 Data File : VX020982.D
 Acq On : 02 Mar 2021 22:10
 Operator : JC/MD
 Sample : M1536-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Mar 03 01:20:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	192522	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	359226	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	355850	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	164557	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	127857	49.17	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.34%
35) Dibromofluoromethane	5.464	113	111489	51.08	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.16%
50) Toluene-d8	8.694	98	450813	52.97	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.94%
62) 4-Bromofluorobenzene	11.118	95	173645	55.18	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.36%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.393	62	9762	3.67	ug/l	95
12) 1,1-Dichloroethene	2.358	96	71255	33.07	ug/l	97
16) Acetone	2.417	43	7119	6.64	ug/l	97
20) Methylene Chloride	2.840	84	926	0.36	ug/l	94
24) 1,1-Dichloroethane	3.670	63	43503	10.43	ug/l	96
27) cis-1,2-Dichloroethene	4.558	96	17317	6.21	ug/l	90
32) 1,1,1-Trichloroethane	5.464	97	8596	2.35	ug/l	98
39) Methylcyclohexane	7.447	83	983	0.23	ug/l #	90
40) Benzene	6.111	78	3948	0.39	ug/l	97
42) 1,2-Dichloroethane	6.164	62	5939	1.75	ug/l	98
44) Trichloroethene	7.188	130	1660	0.62	ug/l	95
49) 1,4-Dioxane	7.712	88	20940	323.65	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030421\
Data File : VX020982.D
Acq On : 02 Mar 2021 22:10
Operator : JC/MD
Sample : M1536-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Mar 03 01:20:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

