

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027298.D
 Acq On : 04 Mar 2022 13:57
 Operator : JC/MD
 Sample : VX0304WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 05 00:43:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	67603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	104166	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	91542	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44971	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	41285	50.849	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.700%
35) Dibromofluoromethane	5.385	113	33284	51.477	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.960%
50) Toluene-d8	8.647	98	119327	49.357	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.720%
62) 4-Bromofluorobenzene	11.079	95	45286	48.799	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.600%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.172	85	11132	16.566	ug/l	98
3) Chloromethane	1.294	50	8984	14.302	ug/l	99
4) Vinyl Chloride	1.380	62	9213	14.725	ug/l	99
5) Bromomethane	1.605	94	5737	20.414	ug/l	98
6) Chloroethane	1.691	64	6215	16.176	ug/l	98
7) Trichlorofluoromethane	1.892	101	18181	18.683	ug/l	100
8) Diethyl Ether	2.136	74	5942	17.543	ug/l	84
9) 1,1,2-Trichlorotrifluo...	2.331	101	12799	18.880	ug/l	95
10) Methyl Iodide	2.459	142	11577	12.718	ug/l	97
11) Tert butyl alcohol	2.959	59	13008	82.344	ug/l	98
12) 1,1-Dichloroethene	2.325	96	11827	17.970	ug/l	80
13) Acrolein	2.239	56	12423	85.081	ug/l	97
14) Allyl chloride	2.666	41	19355	16.703	ug/l	97
15) Acrylonitrile	3.062	53	36634	90.374	ug/l	98
16) Acetone	2.379	43	31385	88.042	ug/l	94
17) Carbon Disulfide	2.514	76	26942	15.743	ug/l	98
18) Methyl Acetate	2.703	43	15907	17.288	ug/l	93
19) Methyl tert-butyl Ether	3.117	73	42402	18.776	ug/l	100
20) Methylene Chloride	2.794	84	13395	18.129	ug/l	92
21) trans-1,2-Dichloroethene	3.093	96	12685	18.073	ug/l	94
22) Diisopropyl ether	3.763	45	40475	17.917	ug/l	95
23) Vinyl Acetate	3.721	43	173306	90.851	ug/l	96
24) 1,1-Dichloroethane	3.617	63	23116	18.102	ug/l	97
25) 2-Butanone	4.556	43	48828	86.009	ug/l	93
26) 2,2-Dichloropropane	4.483	77	17772	19.184	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	14876	18.276	ug/l	94
28) Bromochloromethane	4.903	49	9253	17.910	ug/l	84
29) Tetrahydrofuran	5.007	42	31408	82.662	ug/l	91
30) Chloroform	5.098	83	24484	18.459	ug/l	96
31) Cyclohexane	5.470	56	20068	16.504	ug/l	89
32) 1,1,1-Trichloroethane	5.385	97	22076	19.226	ug/l	98
36) 1,1-Dichloropropene	5.696	75	17810	18.874	ug/l	97
37) Ethyl Acetate	4.721	43	19050	18.387	ug/l	98
38) Carbon Tetrachloride	5.678	117	19730	20.047	ug/l	99
39) Methylcyclohexane	7.385	83	21814	18.584	ug/l	93
40) Benzene	6.043	78	51723	18.743	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027298.D
 Acq On : 04 Mar 2022 13:57
 Operator : JC/MD
 Sample : VX0304WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 05 00:43:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	10108	17.785	ug/l	93
42) 1,2-Dichloroethane	6.086	62	19918	20.346	ug/l	98
43) Isopropyl Acetate	6.342	43	29489	18.502	ug/l	96
44) Trichloroethene	7.123	130	15491	18.766	ug/l	92
45) 1,2-Dichloropropane	7.433	63	12873	19.092	ug/l	99
46) Dibromomethane	7.580	93	9665	19.957	ug/l	93
47) Bromodichloromethane	7.824	83	18505	19.577	ug/l	99
48) Methyl methacrylate	7.696	41	14489	18.781	ug/l	95
49) 1,4-Dioxane	7.659	88	6052	347.004	ug/l	94
51) 4-Methyl-2-Pentanone	8.573	43	95323	92.018	ug/l	95
52) Toluene	8.720	92	33344	18.870	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	18005	18.805	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	21411	20.292	ug/l #	83
55) 1,1,2-Trichloroethane	9.153	97	14001	20.005	ug/l	97
56) Ethyl methacrylate	9.116	69	20822	19.109	ug/l	92
57) 1,3-Dichloropropane	9.311	76	22819	19.603	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	47688	90.797	ug/l	94
59) 2-Hexanone	9.433	43	70999	89.673	ug/l	94
60) Dibromochloromethane	9.518	129	14801	20.386	ug/l	99
61) 1,2-Dibromoethane	9.610	107	14373	19.284	ug/l	99
64) Tetrachloroethene	9.275	164	16047	19.375	ug/l	86
65) Chlorobenzene	10.079	112	36297	19.451	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	14269	21.086	ug/l	99
67) Ethyl Benzene	10.195	91	64598	19.720	ug/l	96
68) m/p-Xylenes	10.305	106	50064	39.754	ug/l	99
69) o-Xylene	10.640	106	24947	19.737	ug/l	96
70) Styrene	10.658	104	41363	19.999	ug/l	97
71) Bromoform	10.799	173	10258	19.694	ug/l #	97
73) Isopropylbenzene	10.963	105	65600	20.386	ug/l	98
74) N-amyl acetate	10.841	43	23287	19.019	ug/l	94
75) 1,1,2,2-Tetrachloroethane	11.213	83	19224	20.226	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	21268	23.479	ug/l	95
77) Bromobenzene	11.201	156	16435	20.476	ug/l	88
78) n-propylbenzene	11.305	91	72569	19.841	ug/l	98
79) 2-Chlorotoluene	11.366	91	43752	19.521	ug/l	93
80) 1,3,5-Trimethylbenzene	11.451	105	55594	20.546	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	5006	19.550	ug/l	97
82) 4-Chlorotoluene	11.457	91	51263	20.022	ug/l	97
83) tert-Butylbenzene	11.713	119	53996	20.522	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	55292	20.543	ug/l	97
85) sec-Butylbenzene	11.890	105	66589	20.184	ug/l	99
86) p-Isopropyltoluene	12.012	119	56634	20.137	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	30474	20.177	ug/l	97
88) 1,4-Dichlorobenzene	12.042	146	30207	19.561	ug/l	99
89) n-Butylbenzene	12.335	91	45968	19.555	ug/l	99
90) Hexachloroethane	12.542	117	8273	18.990	ug/l	92
91) 1,2-Dichlorobenzene	12.335	146	30246	20.524	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	4280	19.396	ug/l	87
93) 1,2,4-Trichlorobenzene	13.591	180	18339	19.721	ug/l	99
94) Hexachlorobutadiene	13.725	225	8077	20.921	ug/l	98
95) Naphthalene	13.774	128	62686	19.837	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	18839	20.211	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
Data File : VX027298.D
Acq On : 04 Mar 2022 13:57
Operator : JC/MD
Sample : VX0304WBSD02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Mar 05 00:43:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:168 Resp: 67603
Ion Ratio Lower Upper
168 100
99 54.9 50.3 75.5



#2
Dichlorodifluoromethane
Concen: 16.566 ug/l
RT: 1.172 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 85 Resp: 11132
Ion Ratio Lower Upper
85 100
87 31.4 16.4 49.1









#3
Chloromethane
Concen: 14.302 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 50 Resp: 8984
Ion Ratio Lower Upper
50 100
52 33.4 26.1 39.1




#4
Vinyl Chloride
Concen: 14.725 ug/l
RT: 1.380 min Scan# 48
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 62 Resp: 9213
Ion Ratio Lower Upper
62 100
64 33.3 26.2 39.2



Abundance Scan 84 (1.599 min): VX027138.D\data.ms (-80)



#5

Bromomethane

Concen: 20.414 ug/l

RT: 1.605 min Scan# 81

Delta R.T. 0.006 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X

ClientSampleId :


Tgt Ion: 94 Resp: 5737

Ion Ratio Lower Upper


94 100

96 93.1 75.9 113.9


Abundance Scan 85 (1.605 min): VX027298.D\data.ms




Abundance



Abundance Scan 85 (1.605 min): VX027298.D\data.ms (-35)



Abundance Scan 97 (1.678 min): VX027138.D\data.ms (-91)



#6

Chloroethane

Concen: 16.176 ug/l

RT: 1.691 min Scan# 99

Delta R.T. 0.013 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57


Tgt Ion: 64 Resp: 6215

Ion Ratio Lower Upper


64 100

66 30.5 25.4 38.2


Abundance Scan 99 (1.691 min): VX027298.D\data.ms




Abundance



Abundance Scan 99 (1.691 min): VX027298.D\data.ms (-48)









#7
 Trichlorofluoromethane
 Concen: 18.683 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX027298.D
 Acq: 04 Mar 2022 13:57




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:101 Resp: 18181
 Ion Ratio Lower Upper
 101 100
 103 64.0 51.3 76.9



#8
 Diethyl Ether
 Concen: 17.543 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX027298.D
 Acq: 04 Mar 2022 13:57

Tgt Ion: 74 Resp: 5942
 Ion Ratio Lower Upper
 74 100
 45 85.2 50.7 152.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.880 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 12799
Ion Ratio Lower Upper
101 100
85 41.3 35.3 52.9
151 81.1 61.4 92.2



#10
Methyl Iodide
Concen: 12.718 ug/l
RT: 2.459 min Scan# 225
Delta R.T. 0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion:142 Resp: 11577
Ion Ratio Lower Upper
142 100
127 45.3 38.1 57.1
141 14.0 11.2 16.8





#11

Tert butyl alcohol

Concen: 82.344 ug/l

RT: 2.959 min Scan# 30

Delta R.T. -0.018 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 13008

Ion Ratio Lower Upper

59 100

57 9.9 7.2 10.8



#12

1,1-Dichloroethene

Concen: 17.970 ug/l

RT: 2.325 min Scan# 203

Delta R.T. 0.006 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57




Tgt Ion: 96 Resp: 11827


Ion Ratio Lower Upper

96 100

61 139.4 138.5 207.7

98 57.3 51.8 77.6





#13

Acrolein

Concen: 85.081 ug/l

RT: 2.239 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 12423

Ion Ratio Lower Upper

56 100

55 68.7 56.9 85.3



#14

Allyl chloride

Concen: 16.703 ug/l

RT: 2.666 min Scan# 259

Delta R.T. 0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57




Tgt Ion: 41 Resp: 19355


Ion Ratio Lower Upper

41 100

39 67.0 54.2 81.4

76 34.8 25.2 37.8









#15
Acrylonitrile
Concen: 90.374 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 53 Resp: 36634
Ion Ratio Lower Upper
53 100
52 81.7 66.6 100.0
51 35.8 29.5 44.3



#16
Acetone
Concen: 88.042 ug/l
RT: 2.379 min Scan# 212
Delta R.T. -0.007 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57


Tgt Ion: 43 Resp: 31385
Ion Ratio Lower Upper
43 100
58 31.9 23.0 34.6








#17
Carbon Disulfide
Concen: 15.743 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 76 Resp: 26942
Ion Ratio Lower Upper
76 100
78 8.7 7.5 11.3




#18
Methyl Acetate
Concen: 17.288 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57



Tgt Ion: 43 Resp: 15907
Ion Ratio Lower Upper
43 100
74 26.3 18.3 27.5









#19
Methyl tert-butyl Ether
Concen: 18.776 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 42402
Ion Ratio Lower Upper
73 100
57 21.8 17.4 26.2



#20
Methylene Chloride
Concen: 18.129 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 84 Resp: 13395
Ion Ratio Lower Upper
84 100
49 113.0 101.5 152.3
51 36.1 31.2 46.8
86 61.9 51.3 76.9









#21
trans-1,2-Dichloroethene
Concen: 18.073 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 12685
Ion Ratio Lower Upper
96 100
61 133.0 112.9 169.3
98 64.9 49.5 74.3



#22
Diisopropyl ether
Concen: 17.917 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 45 Resp: 40475
Ion Ratio Lower Upper
45 100
43 80.6 60.6 91.0
87 28.5 22.0 33.0
59 13.0 9.0 13.4









#23
 Vinyl Acetate
 Concen: 90.851 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX027298.D
 Acq: 04 Mar 2022 13:57




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 173306
 Ion Ratio Lower Upper
 43 100
 86 11.5 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 18.102 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX027298.D
 Acq: 04 Mar 2022 13:57

Tgt Ion: 63 Resp: 23116
 Ion Ratio Lower Upper
 63 100
 98 8.9 3.7 11.1
 100 5.1 2.4 7.2









#25
2-Butanone
Concen: 86.009 ug/l
RT: 4.556 min Scan# 51
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 48828
Ion Ratio Lower Upper
43 100
72 29.1 20.4 30.6



#26
2,2-Dichloropropane
Concen: 19.184 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 77 Resp: 17772
Ion Ratio Lower Upper
77 100
97 22.6 11.0 33.0









#27
cis-1,2-Dichloroethene
Concen: 18.276 ug/l
RT: 4.489 min Scan# 51
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 14876
Ion Ratio Lower Upper
96 100
61 139.5 0.0 296.6
98 66.7 0.0 128.8



#28
Bromochloromethane
Concen: 17.910 ug/l
RT: 4.903 min Scan# 626
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 49 Resp: 9253
Ion Ratio Lower Upper
49 100
129 1.5 0.0 4.2
130 92.0 62.4 93.6





#29

Tetrahydrofuran

Concen: 82.662 ug/l

RT: 5.007 min Scan# 643

Delta R.T. 0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 31408

Ion Ratio Lower Upper

42 100

72 49.1 34.6 52.0

71 44.9 31.8 47.8



#30

Chloroform

Concen: 18.459 ug/l

RT: 5.098 min Scan# 658

Delta R.T. 0.000 min

Lab File: VX027298.D




Acq: 04 Mar 2022 13:57


Tgt Ion: 83 Resp: 24484

Ion Ratio Lower Upper

83 100


85 67.3 51.6 77.4








#31
Cyclohexane
Concen: 16.504 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 56 Resp: 20068




Ion	Ratio	Lower	Upper
56	100		
69	30.5	25.8	38.6
84	99.1	68.9	103.3




#32
1,1,1-Trichloroethane
Concen: 19.226 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 97 Resp: 22076

Ion	Ratio	Lower	Upper
97	100		
99	63.1	51.3	76.9
61	41.5	35.0	52.6





#33

1,2-Dichloroethane-d4

Concen: 50.849 ug/l

RT: 5.958 min Scan# 79

Delta R.T. -0.006 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 41285

Ion Ratio Lower Upper

65 100

67 50.5 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57




Tgt Ion:114 Resp: 104166


Ion Ratio Lower Upper

114 100

63 19.0 0.0 42.8


88 15.9 0.0 33.0








#35
Dibromofluoromethane
Concen: 51.477 ug/l
RT: 5.385 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:113 Resp: 33284
Ion Ratio Lower Upper
113 100
111 104.0 81.2 121.8
192 20.5 15.0 22.4




#36
1,1-Dichloropropene
Concen: 18.874 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57



Tgt Ion: 75 Resp: 17810
Ion Ratio Lower Upper
75 100
110 37.4 17.2 51.6
77 30.4 24.6 36.8





#37

Ethyl Acetate

Concen: 18.387 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.000 min

Lab File: VX027298.D





Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:	43	Resp:	19050
Ion	Ratio	Lower	Upper
43	100		
61	13.8	10.6	16.0
70	12.0	8.5	12.7



#38

Carbon Tetrachloride

Concen: 20.047 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.006 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Tgt Ion:	117	Resp:	19730
Ion	Ratio	Lower	Upper
117	100		
119	93.7	74.7	112.1
121	30.4	25.4	38.0





#39

Methylcyclohexane

Concen: 18.584 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 21814

Ion Ratio Lower Upper

83 100

55 70.2 63.6 95.4

98 47.4 38.4 57.6



#40

Benzene

Concen: 18.743 ug/l

RT: 6.043 min Scan# 813

Delta R.T. 0.000 min

Lab File: VX027298.D




Acq: 04 Mar 2022 13:57


Tgt Ion: 78 Resp: 51723

Ion Ratio Lower Upper

78 100


77 24.7 18.9 28.3








#41
Methacrylonitrile
Concen: 17.785 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41 Resp: 10108
Ion Ratio Lower Upper
41 100
39 57.9 45.4 68.0
67 75.1 53.2 79.8
52 32.4 23.0 34.6



#42
1,2-Dichloroethane
Concen: 20.346 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 62 Resp: 19918
Ion Ratio Lower Upper
62 100
98 9.5 0.0 17.8









#43
Isopropyl Acetate
Concen: 18.502 ug/l
RT: 6.342 min Scan# 80
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 29489
Ion Ratio Lower Upper
43 100
61 21.0 15.4 23.0
87 14.1 10.2 15.2



#44
Trichloroethene
Concen: 18.766 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion:130 Resp: 15491
Ion Ratio Lower Upper
130 100
95 93.5 0.0 202.6





#45

1,2-Dichloropropane

Concen: 19.092 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 12873

Ion Ratio Lower Upper

63 100

65 30.0 23.8 35.6



#46

Dibromomethane

Concen: 19.957 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57




Tgt Ion: 93 Resp: 9665


Ion Ratio Lower Upper

93 100

95 82.4 66.1 99.1

174 110.8 78.7 118.1









#47
Bromodichloromethane
Concen: 19.577 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 18505
Ion Ratio Lower Upper
83 100
85 63.7 51.4 77.0
127 9.3 6.9 10.3



#48
Methyl methacrylate
Concen: 18.781 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 41 Resp: 14489
Ion Ratio Lower Upper
41 100
69 85.5 62.2 93.4
39 56.5 45.4 68.0









#49
1,4-Dioxane
Concen: 347.004 ug/l
RT: 7.659 min Scan# 1079
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 88 Resp: 6052
Ion Ratio Lower Upper
88 100
43 33.0 29.9 44.9
58 67.4 57.8 86.6



#50
Toluene-d8
Concen: 49.357 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 98 Resp: 119327
Ion Ratio Lower Upper
98 100
100 66.9 53.2 79.8









#51
4-Methyl-2-Pentanone
Concen: 92.018 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.007 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 95323
Ion Ratio Lower Upper
43 100
58 41.0 30.5 45.7



#52
Toluene
Concen: 18.870 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 92 Resp: 33344
Ion Ratio Lower Upper
92 100
91 172.1 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 18.805 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 18005

Ion Ratio Lower Upper

75 100

77 32.8 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 20.292 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57




Tgt Ion: 75 Resp: 21411


Ion Ratio Lower Upper

75 100

77 25.6 26.2 39.2#

39 40.6 43.5 65.3#









#55
1,1,2-Trichloroethane
Concen: 20.005 ug/l
RT: 9.153 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 97 Resp: 14001




Ion	Ratio	Lower	Upper
97	100		
83	82.1	68.4	102.6
85	53.2	44.0	66.0
99	63.5	49.8	74.6




#56
Ethyl methacrylate
Concen: 19.109 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 69 Resp: 20822

Ion	Ratio	Lower	Upper
69	100		
41	64.4	57.2	85.8
39	34.9	30.9	46.3









#57
1,3-Dichloropropane
Concen: 19.603 ug/l
RT: 9.311 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 22819
Ion Ratio Lower Upper
76 100
78 32.3 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 90.797 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 63 Resp: 47688
Ion Ratio Lower Upper
63 100
106 31.9 23.0 34.6





#59

2-Hexanone

Concen: 89.673 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 70999

Ion Ratio Lower Upper

43 100

58 56.9 26.4 79.2



#60

Dibromochloromethane

Concen: 20.386 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.007 min

Lab File: VX027298.D




Acq: 04 Mar 2022 13:57


Tgt Ion: 129 Resp: 14801

Ion Ratio Lower Upper

129 100

127 76.1 38.5 115.5





#61

1,2-Dibromoethane

Concen: 19.284 ug/l

RT: 9.610 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:107 Resp: 14373

Ion Ratio Lower Upper

107 100

109 95.3 75.5 113.3



#62

4-Bromofluorobenzene

Concen: 48.799 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.006 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57




Tgt Ion: 95 Resp: 45286


Ion Ratio Lower Upper

95 100

174 81.7 0.0 148.6

176 78.6 0.0 141.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
117	100		
82	53.7	45.2	67.8
119	31.9	26.2	39.2



#64
Tetrachloroethene
Concen: 19.375 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion	Ratio	Lower	Upper
164	100		
166	146.2	101.8	152.6
129	88.5	83.3	124.9
131	87.2	78.0	117.0









#65
Chlorobenzene
Concen: 19.451 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 36297
Ion Ratio Lower Upper
112 100
114 31.3 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 21.086 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion:131 Resp: 14269
Ion Ratio Lower Upper
131 100
133 95.7 47.9 143.8
119 61.2 31.7 95.1





#67

Ethyl Benzene

Concen: 19.720 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 64598

Ion Ratio Lower Upper

91 100

106 32.9 24.7 37.1



#68

m/p-Xylenes

Concen: 39.754 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. -0.000 min

Lab File: VX027298.D




Acq: 04 Mar 2022 13:57


Tgt Ion:106 Resp: 50064

Ion Ratio Lower Upper

106 100

91 204.3 165.3 247.9









#69
o-Xylene
Concen: 19.737 ug/l
RT: 10.640 min Scan# 1568
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 24947
Ion Ratio Lower Upper
106 100
91 213.1 109.7 329.0



#70
Styrene
Concen: 19.999 ug/l
RT: 10.658 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion:104 Resp: 41363
Ion Ratio Lower Upper
104 100
78 48.8 42.5 63.7
103 54.1 43.3 64.9









#71
Bromoform
Concen: 19.694 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:173 Resp: 10258
Ion Ratio Lower Upper
173 100
175 50.5 24.3 72.9
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion:152 Resp: 44971
Ion Ratio Lower Upper
152 100
115 68.0 44.2 132.6
150 164.0 0.0 348.2









#73
Isopropylbenzene
Concen: 20.386 ug/l
RT: 10.963 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 65600
Ion Ratio Lower Upper
105 100
120 26.9 13.1 39.1



#74
N-ethyl acetate
Concen: 19.019 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 43 Resp: 23287
Ion Ratio Lower Upper
43 100
70 46.1 32.9 49.3
55 24.9 18.3 27.5
61 25.0 18.9 28.3









#75
 1,1,2,2-Tetrachloroethane
 Concen: 20.226 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX027298.D
 Acq: 04 Mar 2022 13:57




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 83 Resp: 19224
 Ion Ratio Lower Upper
 83 100
 131 11.1 5.3 15.9
 85 64.1 32.0 96.2



#76
 1,2,3-Trichloropropane
 Concen: 23.479 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.006 min
 Lab File: VX027298.D
 Acq: 04 Mar 2022 13:57


Tgt Ion: 75 Resp: 21268
 Ion Ratio Lower Upper
 75 100
 77 33.5 18.3 54.8








#77
Bromobenzene
Concen: 20.476 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:156 Resp: 16435
Ion Ratio Lower Upper
156 100
77 144.3 84.5 253.4
158 95.7 48.1 144.4



#78
n-propylbenzene
Concen: 19.841 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 91 Resp: 72569
Ion Ratio Lower Upper
91 100
120 24.1 11.6 34.7









#79
2-Chlorotoluene
Concen: 19.521 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 43752
Ion Ratio Lower Upper
91 100
126 36.7 16.4 49.2



#80
1,3,5-Trimethylbenzene
Concen: 20.546 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57


Tgt Ion: 105 Resp: 55594
Ion Ratio Lower Upper
105 100
120 48.6 24.0 72.0








#81
trans-1,4-Dichloro-2-butene
Concen: 19.550 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 75 Resp: 5006
Ion Ratio Lower Upper
75 100
53 112.5 94.0 141.0
89 46.7 37.0 55.4



#82
4-Chlorotoluene
Concen: 20.022 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 91 Resp: 51263
Ion Ratio Lower Upper
91 100
126 29.9 14.1 42.4





#83

tert-Butylbenzene

Concen: 20.522 ug/l

RT: 11.713 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX027298.D





Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:	119	Resp:	53996
Ion Ratio	Lower	Upper	
119	100		
91	58.4	31.4	94.0
134	23.7	11.8	35.3



#84

1,2,4-Trimethylbenzene

Concen: 20.543 ug/l




RT: 11.756 min Scan# 1750


Delta R.T. -0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Tgt Ion:	105	Resp:	55292
Ion Ratio	Lower	Upper	
105	100		
120	45.9	21.9	65.5









#85
 sec-Butylbenzene
 Concen: 20.184 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX027298.D
 Acq: 04 Mar 2022 13:57




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 66589
 Ion Ratio Lower Upper
 105 100
 134 20.5 10.1 30.1



#86
 p-Isopropyltoluene
 Concen: 20.137 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX027298.D
 Acq: 04 Mar 2022 13:57

Tgt Ion:119 Resp: 56634
 Ion Ratio Lower Upper
 119 100
 134 26.9 13.0 39.0
 91 23.9 12.6 37.6





#87

1,3-Dichlorobenzene

Concen: 20.177 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 30474

Ion Ratio Lower Upper

146 100

111 39.0 21.3 63.7

148 62.2 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 19.561 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57




Tgt Ion:146 Resp: 30207


Ion Ratio Lower Upper

146 100

111 40.0 20.2 60.6

148 64.2 31.4 94.0









#89
n-Butylbenzene
Concen: 19.555 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 91 Resp: 45968




Ion	Ratio	Lower	Upper
91	100		
92	54.0	27.2	81.6
134	27.7	13.1	39.1




#90
Hexachloroethane
Concen: 18.990 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion: 117 Resp: 8273

Ion	Ratio	Lower	Upper
117	100		
201	77.5	35.6	106.8





#91

1,2-Dichlorobenzene

Concen: 20.524 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 30246

Ion Ratio Lower Upper

146 100

111 42.1 21.8 65.3

148 63.7 32.0 96.2



#92

1,2-Dibromo-3-Chloropropane

Concen: 19.396 ug/l

RT: 12.945 min Scan# 1945

Delta R.T. -0.000 min

Lab File: VX027298.D

Acq: 04 Mar 2022 13:57




Tgt Ion: 75 Resp: 4280


Ion Ratio Lower Upper

75 100

155 89.9 40.0 119.9


157 117.7 51.6 154.8








#93
1,2,4-Trichlorobenzene
Concen: 19.721 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:180 Resp: 18339
Ion Ratio Lower Upper
180 100
182 95.0 46.7 140.0
145 32.6 16.3 48.8



#94
Hexachlorobutadiene
Concen: 20.921 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57


Tgt Ion:225 Resp: 8077
Ion Ratio Lower Upper
225 100
223 65.5 32.0 96.0
227 65.7 32.3 96.9








#95
Naphthalene
Concen: 19.837 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.006 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 62686
Ion Ratio Lower Upper
128 100
127 12.9 10.3 15.5
129 11.0 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 20.211 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX027298.D
Acq: 04 Mar 2022 13:57

Tgt Ion:180 Resp: 18839
Ion Ratio Lower Upper
180 100
182 93.1 47.2 141.6
145 33.8 17.5 52.5

