

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027292.D
 Acq On : 04 Mar 2022 11:25
 Operator : JC/MD
 Sample : VX0304WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0304WBL01

Quant Time: Mar 05 01:48:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030322W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 04 01:00:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	12323	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	67252	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	58213	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	26404	30.950	ug/l	-0.01
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.167%
60) 4-Bromofluorobenzene	11.079	95	25876	27.438	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.467%
63) Toluene-d8	8.647	98	77044	30.274	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
Data File : VX027292.D
Acq On : 04 Mar 2022 11:25
Operator : JC/MD
Sample : VX0304WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0304WBL01

Quant Time: Mar 05 01:48:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030322W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 04 01:00:22 2022
Response via : Initial Calibration

