

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027296.D
 Acq On : 04 Mar 2022 13:07
 Operator : JC/MD
 Sample : N1799-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-038-IDWGW-20220303

Quant Time: Mar 05 01:49:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030322W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 04 01:00:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	10799	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	58552	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	50421	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	23145	30.958	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.200%
60) 4-Bromofluorobenzene	11.079	95	22375	27.393	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.300%
63) Toluene-d8	8.653	98	66197	30.032	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.100%
Target Compounds						
37) Trichloroethene	7.129	130	5200	6.580	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
Data File : VX027296.D
Acq On : 04 Mar 2022 13:07
Operator : JC/MD
Sample : N1799-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-038-IDWGW-20220303

Quant Time: Mar 05 01:49:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030322W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 04 01:00:22 2022
Response via : Initial Calibration

